

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091522\
 Data File : VN074464.D
 Acq On : 15 Sep 2022 16:20
 Operator : JC\MD
 Sample : N4471-15DL 200X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-06-1.5-11.5-083122DL

Quant Time: Sep 20 07:24:52 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091522W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 15 19:20:33 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	449470	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.899	114	871349	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.687	117	854896	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.616	152	363118	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	323217	42.382	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	84.760%		
35) Dibromofluoromethane	7.946	113	273323	47.221	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	94.440%		
50) Toluene-d8	10.381	98	994158	44.411	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	88.820%		
62) 4-Bromofluorobenzene	12.675	95	403101	52.577	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	105.160%		
Target Compounds						Qvalue
9) 1,1,2-Trichlorotrifluo...	4.140	101	247574	41.155	ug/l	90
27) cis-1,2-Dichloroethene	7.246	96	58760	7.060	ug/l	100
44) Trichloroethene	9.151	130	2175338	317.318	ug/l	86
64) Tetrachloroethene	10.916	164	10970	1.870	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091522\
Data File : VN074464.D
Acq On : 15 Sep 2022 16:20
Operator : JC\MD
Sample : N4471-15DL 200X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-06-1.5-11.5-083122DL

Quant Time: Sep 20 07:24:52 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091522W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 15 19:20:33 2022
Response via : Initial Calibration

