

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091522\
 Data File : VN074466.D
 Acq On : 15 Sep 2022 17:09
 Operator : JC\MD
 Sample : N4471-10RE
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-14A-15-20-083122RE

Quant Time: Sep 16 01:46:09 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091522W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 15 19:20:33 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	444747	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	846260	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	842210	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.616	152	374175	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	494306	65.504	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	131.000%#
35) Dibromofluoromethane	7.951	113	386953	68.834	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	137.660%#
50) Toluene-d8	10.380	98	1435889	66.046	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	132.100%#
62) 4-Bromofluorobenzene	12.674	95	521703	70.064	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	140.120%#
Target Compounds						Qvalue
20) Methylene Chloride	5.022	84	12624	1.742	ug/l	# 91
30) Chloroform	7.739	83	8817	0.680	ug/l	90
44) Trichloroethene	9.145	130	3641	0.547	ug/l	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091522\
Data File : VN074466.D
Acq On : 15 Sep 2022 17:09
Operator : JC\MD
Sample : N4471-10RE
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-14A-15-20-083122RE

Quant Time: Sep 16 01:46:09 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091522W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 15 19:20:33 2022
Response via : Initial Calibration

