

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091522\
 Data File : VN074471.D
 Acq On : 15 Sep 2022 19:11
 Operator : JC\MD
 Sample : N4600-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 OH-SOIL-0913

Quant Time: Sep 16 01:46:31 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091522W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 15 19:20:33 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	432598	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	850758	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	856313	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.616	152	367478	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	315515	42.985	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	85.980%
35) Dibromofluoromethane	7.951	113	259663	45.946	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.900%
50) Toluene-d8	10.381	98	979568	44.819	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	89.640%
62) 4-Bromofluorobenzene	12.675	95	393474	52.564	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	105.120%
Target Compounds						
					Qvalue	
16) Acetone	4.222	43	85177	18.135	ug/l	97
20) Methylene Chloride	5.022	84	29845	4.190	ug/l	90
37) Ethyl Acetate	7.340	43	119191	7.715	ug/l	96
43) Isopropyl Acetate	8.492	43	627923	25.949	ug/l #	80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091522\
Data File : VN074471.D
Acq On : 15 Sep 2022 19:11
Operator : JC\MD
Sample : N4600-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
OH-SOIL-0913

Quant Time: Sep 16 01:46:31 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091522W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 15 19:20:33 2022
Response via : Initial Calibration

