

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091522\
 Data File : VN074479.D
 Acq On : 15 Sep 2022 23:14
 Operator : JC\MD
 Sample : VN0915WBL02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0915WBL02

Quant Time: Sep 16 01:47:53 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091522W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 15 19:20:33 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	421446	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	800075	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	780195	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.616	152	339399	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	304567	42.592	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	85.180%
35) Dibromofluoromethane	7.951	113	258911	48.716	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.440%
50) Toluene-d8	10.381	98	945421	45.997	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	92.000%
62) 4-Bromofluorobenzene	12.675	95	379040	53.843	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	107.680%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091522\
Data File : VN074479.D
Acq On : 15 Sep 2022 23:14
Operator : JC\MD
Sample : VN0915WBL02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 31 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0915WBL02

Quant Time: Sep 16 01:47:53 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091522W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 15 19:20:33 2022
Response via : Initial Calibration

