

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091522\
 Data File : VN074483.D
 Acq On : 16 Sep 2022 00:51
 Operator : JC\MD
 Sample : N4511-01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 S1-TB-20220902

Quant Time: Sep 16 01:48:44 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091522W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 15 19:20:33 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	395406	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	774041	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	767028	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.616	152	327154	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	305794	45.580	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	91.160%
35) Dibromofluoromethane	7.951	113	254098	49.418	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.840%
50) Toluene-d8	10.380	98	932016	46.870	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	93.740%
62) 4-Bromofluorobenzene	12.674	95	368468	54.102	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	108.200%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091522\
Data File : VN074483.D
Acq On : 16 Sep 2022 00:51
Operator : JC\MD
Sample : N4511-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 35 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
S1-TB-20220902

Quant Time: Sep 16 01:48:44 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091522W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 15 19:20:33 2022
Response via : Initial Calibration

