

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091522\
 Data File : VN074484.D
 Acq On : 16 Sep 2022 01:15
 Operator : JC\MD
 Sample : N4630-15
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-17-4-3-GR

Quant Time: Sep 16 07:00:21 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091522W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 15 19:20:33 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	397208	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	777599	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	817452	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.616	152	340398	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.375	65	298075	44.227	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	88.460%
35) Dibromofluoromethane	7.951	113	247049	47.827	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.660%
50) Toluene-d8	10.381	98	918522	45.980	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	91.960%
62) 4-Bromofluorobenzene	12.674	95	378310	55.293	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	110.580%
Target Compounds						
					Qvalue	
16) Acetone	4.222	43	79098	18.341	ug/l	89
20) Methylene Chloride	5.010	84	21927	3.359	ug/l #	91
37) Ethyl Acetate	7.334	43	102820	7.282	ug/l	95
43) Isopropyl Acetate	8.492	43	446598	20.192	ug/l #	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091522\
Data File : VN074484.D
Acq On : 16 Sep 2022 01:15
Operator : JC\MD
Sample : N4630-15
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 36 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-17-4-3-GR

Quant Time: Sep 16 07:00:21 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091522W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 15 19:20:33 2022
Response via : Initial Calibration

