

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091522\
 Data File : VN074485.D
 Acq On : 16 Sep 2022 01:40
 Operator : JC\MD
 Sample : N4630-05
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-17-4-1-GR

Quant Time: Sep 16 07:00:38 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091522W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 15 19:20:33 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	420908	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	825858	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	831641	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.616	152	365265	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	317715	44.487	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	88.980%
35) Dibromofluoromethane	7.951	113	255543	46.581	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.160%
50) Toluene-d8	10.380	98	962698	45.375	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	90.740%
62) 4-Bromofluorobenzene	12.674	95	388522	53.467	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	106.940%
Target Compounds						
					Qvalue	
16) Acetone	4.216	43	84605	18.513	ug/l	90
20) Methylene Chloride	5.010	84	21872	3.164	ug/l #	88
37) Ethyl Acetate	7.334	43	100781	6.720	ug/l #	95
43) Isopropyl Acetate	8.492	43	477201	20.315	ug/l #	82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091522\
Data File : VN074485.D
Acq On : 16 Sep 2022 01:40
Operator : JC\MD
Sample : N4630-05
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 37 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-17-4-1-GR

Quant Time: Sep 16 07:00:38 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091522W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 15 19:20:33 2022
Response via : Initial Calibration

