

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091522\
 Data File : VN074488.D
 Acq On : 16 Sep 2022 02:52
 Operator : JC\MD
 Sample : N4630-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-17-4-1

Quant Time: Sep 16 07:01:13 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091522W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 15 19:20:33 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	396343	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	768914	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	795698	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.616	152	371225	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	299471	44.532	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	89.060%		
35) Dibromofluoromethane	7.951	113	243147	47.603	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	95.200%		
50) Toluene-d8	10.381	98	922862	46.719	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	93.440%		
62) 4-Bromofluorobenzene	12.674	95	389835	57.621	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	115.240%		
Target Compounds					Qvalue	
16) Acetone	4.216	43	124977	29.042	ug/l	90
18) Methyl Acetate	4.769	43	16815	1.258	ug/l #	73
20) Methylene Chloride	5.004	84	25255	3.873	ug/l #	93
25) 2-Butanone	7.269	43	35998	5.038	ug/l #	88
37) Ethyl Acetate	7.334	43	99895	7.155	ug/l #	96
43) Isopropyl Acetate	8.492	43	451089	20.626	ug/l #	82
52) Toluene	10.445	92	33781	2.061	ug/l	90
95) Naphthalene	15.445	128	46988	3.021	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091522\
Data File : VN074488.D
Acq On : 16 Sep 2022 02:52
Operator : JC\MD
Sample : N4630-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-17-4-1

Quant Time: Sep 16 07:01:13 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091522W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 15 19:20:33 2022
Response via : Initial Calibration

