

Data Path : W:\HPCHEM1\MSVOA N\DATA\VN091715\
 Data File : VN027254.D
 Acq On : 17 Sep 2015 14:45
 Operator : MD\FY
 Sample : G3663-01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 GW-1

Manual Integrations
 APPROVED

MMDadoda
 9/18/2015 12:23:02 PM

Quant Time: Sep 18 12:12:02 2015
 Quant Method : W:\HPCHEM1\MSVOA_N\METHODS\82N091515W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 15 21:12:09 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.75	168	750318	50.00	ug/l	0.00
35) 1,4-Difluorobenzene	8.68	114	1331542	50.00	ug/l	0.00
64) Chlorobenzene-d5	11.52	117	1131544	50.00	ug/l	0.00
73) 1,4-Dichlorobenzene-d4	13.47	152	390771	50.00	ug/l	0.00

System Monitoring Compounds

34) 1,2-Dichloroethane-d4	8.12	65	746197	48.54	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.08%	
36) Dibromofluoromethane	7.68	113	480280	52.03	ug/l	0.00
Spiked Amount	50.000		Recovery	=	104.06%	
51) Toluene-d8	10.19	98	1950808	53.98	ug/l	0.00
Spiked Amount	50.000		Recovery	=	107.96%	
63) 4-Bromofluorobenzene	12.52	95	647593	50.49	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.98%	

Target Compounds

	R.T.	QIon	Response	Conc	Units	Qvalue
20) Methyl tert-butyl Ether	5.11	73	36660m	0.95	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA N\DATA\VN091715\
Data File : VN027254.D
Acq On : 17 Sep 2015 14:45
Operator : MD\FY
Sample : G3663-01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
GW-1

Manual Integrations
APPROVED

MMDadoda
9/18/2015 12:23:02 PM

Quant Time: Sep 18 12:12:02 2015
Quant Method : W:\HPCHEM1\MSVOA_N\METHODS\82N091515W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 15 21:12:09 2015
Response via : Initial Calibration

