

Data Path : W:\HPCHEM1\MSVOA N\DATA\VN091715\  
 Data File : VN027258.D  
 Acq On : 17 Sep 2015 16:36  
 Operator : MD\FY  
 Sample : G3704-02  
 Misc : 5.00mL/MSVOA N/WATER  
 ALS Vial : 17 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 MW-3-9-15-15

Quant Time: Sep 18 03:19:30 2015  
 Quant Method : W:\HPCHEM1\MSVOA\_N\METHODS\82N091515W.M  
 Quant Title : SW846 8260  
 QLast Update : Tue Sep 15 21:12:09 2015  
 Response via : Initial Calibration

| Internal Standards          | R.T.  | QIon | Response | Conc  | Units   | Dev(Min) |
|-----------------------------|-------|------|----------|-------|---------|----------|
| 1) Pentafluorobenzene       | 7.75  | 168  | 770872   | 50.00 | ug/l    | 0.00     |
| 35) 1,4-Difluorobenzene     | 8.68  | 114  | 1382501  | 50.00 | ug/l    | 0.00     |
| 64) Chlorobenzene-d5        | 11.52 | 117  | 1159104  | 50.00 | ug/l    | 0.00     |
| 73) 1,4-Dichlorobenzene-d4  | 13.47 | 152  | 405533   | 50.00 | ug/l    | 0.00     |
| System Monitoring Compounds |       |      |          |       |         |          |
| 34) 1,2-Dichloroethane-d4   | 8.12  | 65   | 772016   | 48.88 | ug/l    | 0.00     |
| Spiked Amount 50.000        |       |      | Recovery | =     | 97.76%  |          |
| 36) Dibromofluoromethane    | 7.68  | 113  | 489712   | 51.10 | ug/l    | 0.00     |
| Spiked Amount 50.000        |       |      | Recovery | =     | 102.20% |          |
| 51) Toluene-d8              | 10.19 | 98   | 2010845  | 53.59 | ug/l    | 0.00     |
| Spiked Amount 50.000        |       |      | Recovery | =     | 107.18% |          |
| 63) 4-Bromofluorobenzene    | 12.52 | 95   | 667894   | 50.15 | ug/l    | 0.00     |
| Spiked Amount 50.000        |       |      | Recovery | =     | 100.30% |          |
| Target Compounds            |       |      |          |       |         |          |
| 31) Chloroform              | 7.46  | 83   | 6415     | 0.29  | ug/l    | 92       |
| 65) Tetrachloroethene       | 10.74 | 164  | 20480    | 2.76  | ug/l    | 91       |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA N\DATA\VN091715\  
Data File : VN027258.D  
Acq On : 17 Sep 2015 16:36  
Operator : MD\FY  
Sample : G3704-02  
Misc : 5.00mL/MSVOA N/WATER  
ALS Vial : 17 Sample Multiplier: 1

Instrument :  
MSVOA\_N  
ClientSampleId :  
MW-3-9-15-15

Quant Time: Sep 18 03:19:30 2015  
Quant Method : W:\HPCHEM1\MSVOA\_N\METHODS\82N091515W.M  
Quant Title : SW846 8260  
QLast Update : Tue Sep 15 21:12:09 2015  
Response via : Initial Calibration

