

Data Path : W:\HPCHEM1\MSVOA N\DATA\VN091715\
 Data File : VN027262.D
 Acq On : 17 Sep 2015 18:26
 Operator : MD\FY
 Sample : G3704-06
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-7-9-15-15

Quant Time: Sep 18 03:38:57 2015
 Quant Method : W:\HPCHEM1\MSVOA_N\METHODS\82N091515W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 15 21:12:09 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.75	168	751238	50.00	ug/l	0.00
35) 1,4-Difluorobenzene	8.68	114	1370562	50.00	ug/l	0.00
64) Chlorobenzene-d5	11.52	117	1146858	50.00	ug/l	0.00
73) 1,4-Dichlorobenzene-d4	13.47	152	402957	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
34) 1,2-Dichloroethane-d4	8.12	65	753574	48.96	ug/l	0.00
Spiked Amount						
			Recovery	=	97.92%	
36) Dibromofluoromethane	7.68	113	488139	51.38	ug/l	0.00
Spiked Amount						
			Recovery	=	102.76%	
51) Toluene-d8	10.19	98	1984580	53.35	ug/l	0.00
Spiked Amount						
			Recovery	=	106.70%	
63) 4-Bromofluorobenzene	12.52	95	661778	50.12	ug/l	0.00
Spiked Amount						
			Recovery	=	100.24%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
65) Tetrachloroethene	10.74	164	3401	0.46	ug/l	# 81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA N\DATA\VN091715\
Data File : VN027262.D
Acq On : 17 Sep 2015 18:26
Operator : MD\FY
Sample : G3704-06
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-7-9-15-15

Quant Time: Sep 18 03:38:57 2015
Quant Method : W:\HPCHEM1\MSVOA_N\METHODS\82N091515W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 15 21:12:09 2015
Response via : Initial Calibration

