

Data Path : W:\HPCHEM1\MSVOA N\DATA\VN091715\  
 Data File : VN027283.D  
 Acq On : 18 Sep 2015 5:00  
 Operator : MD\FY  
 Sample : G3706-06  
 Misc : 5.00mL/MSVOA N/WATER  
 ALS Vial : 42 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 MW-7

Quant Time: Sep 18 12:03:16 2015  
 Quant Method : W:\HPCHEM1\MSVOA\_N\METHODS\82N091515W.M  
 Quant Title : SW846 8260  
 QLast Update : Tue Sep 15 21:12:09 2015  
 Response via : Initial Calibration

| Internal Standards         | R.T.  | QIon | Response | Conc  | Units | Dev(Min) |
|----------------------------|-------|------|----------|-------|-------|----------|
| 1) Pentafluorobenzene      | 7.75  | 168  | 751222   | 50.00 | ug/l  | 0.00     |
| 35) 1,4-Difluorobenzene    | 8.68  | 114  | 1369064  | 50.00 | ug/l  | 0.00     |
| 64) Chlorobenzene-d5       | 11.52 | 117  | 1143467  | 50.00 | ug/l  | 0.00     |
| 73) 1,4-Dichlorobenzene-d4 | 13.47 | 152  | 412120   | 50.00 | ug/l  | 0.00     |

| System Monitoring Compounds | R.T.  | QIon | Response | Conc  | Units   | Dev(Min) |
|-----------------------------|-------|------|----------|-------|---------|----------|
| 34) 1,2-Dichloroethane-d4   | 8.12  | 65   | 762780   | 49.56 | ug/l    | 0.00     |
| Spiked Amount               |       |      | Recovery | =     | 99.12%  |          |
| 36) Dibromofluoromethane    | 7.68  | 113  | 484435   | 51.05 | ug/l    | 0.00     |
| Spiked Amount               |       |      | Recovery | =     | 102.10% |          |
| 51) Toluene-d8              | 10.19 | 98   | 1990285  | 53.56 | ug/l    | 0.00     |
| Spiked Amount               |       |      | Recovery | =     | 107.12% |          |
| 63) 4-Bromofluorobenzene    | 12.52 | 95   | 670283   | 50.82 | ug/l    | 0.00     |
| Spiked Amount               |       |      | Recovery | =     | 101.64% |          |

| Target Compounds            | R.T. | QIon | Response | Conc | Units | Qvalue |
|-----------------------------|------|------|----------|------|-------|--------|
| 20) Methyl tert-butyl Ether | 5.12 | 73   | 24934    | 0.64 | ug/l  | # 90   |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA N\DATA\VN091715\  
Data File : VN027283.D  
Acq On : 18 Sep 2015 5:00  
Operator : MD\FY  
Sample : G3706-06  
Misc : 5.00mL/MSVOA N/WATER  
ALS Vial : 42 Sample Multiplier: 1

Instrument :  
MSVOA\_N  
ClientSampleId :  
MW-7

Quant Time: Sep 18 12:03:16 2015  
Quant Method : W:\HPCHEM1\MSVOA\_N\METHODS\82N091515W.M  
Quant Title : SW846 8260  
QLast Update : Tue Sep 15 21:12:09 2015  
Response via : Initial Calibration

