

Data Path : W:\HPCHEM1\MSVOA N\DATA\VN091715\
 Data File : VN027284.D
 Acq On : 18 Sep 2015 5:28
 Operator : MD\FY
 Sample : G3706-07
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-8

Quant Time: Sep 18 08:03:44 2015
 Quant Method : W:\HPCHEM1\MSVOA_N\METHODS\82N091515W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 15 21:12:09 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.75	168	729636	50.00	ug/l	0.00
35) 1,4-Difluorobenzene	8.68	114	1287739	50.00	ug/l	0.00
64) Chlorobenzene-d5	11.52	117	1112139	50.00	ug/l	0.00
73) 1,4-Dichlorobenzene-d4	13.47	152	411743	50.00	ug/l	0.00

System Monitoring Compounds						
34) 1,2-Dichloroethane-d4	8.12	65	702085	46.96	ug/l	0.00
Spiked Amount	50.000		Recovery	=	93.92%	
36) Dibromofluoromethane	7.68	113	464598	52.05	ug/l	0.00
Spiked Amount	50.000		Recovery	=	104.10%	
51) Toluene-d8	10.19	98	1923483	55.03	ug/l	0.00
Spiked Amount	50.000		Recovery	=	110.06%	
63) 4-Bromofluorobenzene	12.52	95	654287	52.74	ug/l	0.00
Spiked Amount	50.000		Recovery	=	105.48%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA N\DATA\VN091715\
Data File : VN027284.D
Acq On : 18 Sep 2015 5:28
Operator : MD\FY
Sample : G3706-07
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 43 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-8

Quant Time: Sep 18 08:03:44 2015
Quant Method : W:\HPCHEM1\MSVOA_N\METHODS\82N091515W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 15 21:12:09 2015
Response via : Initial Calibration

