

Data Path : W:\HPCHEM1\MSVOA N\DATA\VN091715\
 Data File : VN027287.D
 Acq On : 18 Sep 2015 6:51
 Operator : MD\FY
 Sample : G3706-10
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 46 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 DUP-01

Quant Time: Sep 18 08:18:05 2015
 Quant Method : W:\HPCHEM1\MSVOA_N\METHODS\82N091515W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 15 21:12:09 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.75	168	749120	50.00	ug/l	0.00
35) 1,4-Difluorobenzene	8.68	114	1336068	50.00	ug/l	0.00
64) Chlorobenzene-d5	11.52	117	1134986	50.00	ug/l	0.00
73) 1,4-Dichlorobenzene-d4	13.47	152	408784	50.00	ug/l	0.00
System Monitoring Compounds						
34) 1,2-Dichloroethane-d4	8.12	65	733769	47.81	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.62%	
36) Dibromofluoromethane	7.68	113	480406	51.87	ug/l	0.00
Spiked Amount	50.000		Recovery	=	103.74%	
51) Toluene-d8	10.19	98	1965894	54.21	ug/l	0.00
Spiked Amount	50.000		Recovery	=	108.42%	
63) 4-Bromofluorobenzene	12.52	95	661428	51.39	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.78%	
Target Compounds						
13) 1,1-Dichloroethene	3.77	96	68508	7.12	ug/l	88
25) 1,1-Dichloroethane	5.93	63	81458	3.44	ug/l	98
33) 1,1,1-Trichloroethane	7.67	97	377160	19.77	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA N\DATA\VN091715\
Data File : VN027287.D
Acq On : 18 Sep 2015 6:51
Operator : MD\FY
Sample : G3706-10
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 46 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
DUP-01

Quant Time: Sep 18 08:18:05 2015
Quant Method : W:\HPCHEM1\MSVOA_N\METHODS\82N091515W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 15 21:12:09 2015
Response via : Initial Calibration

