

Data Path : W:\HPCHEM1\MSVOA N\DATA\VN092115\
 Data File : VN027362.D
 Acq On : 22 Sep 2015 6:39
 Operator : MD\FY
 Sample : G3754-07
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-9B

Quant Time: Sep 22 08:22:33 2015
 Quant Method : W:\HPCHEM1\MSVOA_N\METHODS\82N092115W.M
 Quant Title : SW846 8260
 QLast Update : Mon Sep 21 13:12:32 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.75	168	712561	50.00	ug/l	0.00
35) 1,4-Difluorobenzene	8.68	114	1285285	50.00	ug/l	0.00
64) Chlorobenzene-d5	11.52	117	1098668	50.00	ug/l	0.00
73) 1,4-Dichlorobenzene-d4	13.47	152	397142	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
34) 1,2-Dichloroethane-d4	8.12	65	758916	54.41	ug/l	0.00
Spiked Amount						
			Recovery	=	108.82%	
36) Dibromofluoromethane	7.68	113	465620	55.52	ug/l	0.00
Spiked Amount						
			Recovery	=	111.04%	
51) Toluene-d8	10.19	98	1895776	58.47	ug/l	0.00
Spiked Amount						
			Recovery	=	116.94%	
63) 4-Bromofluorobenzene	12.52	95	655348	55.64	ug/l	0.00
Spiked Amount						
			Recovery	=	111.28%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
17) Acetone	3.84	43	35831	9.44	ug/l	96
41) Benzene	8.14	78	11443	0.28	ug/l #	90
53) Toluene	10.26	92	16541	0.70	ug/l	98
68) Ethyl Benzene	11.62	91	43386	0.98	ug/l	98
69) m/p-Xylenes	11.72	106	21669	1.41	ug/l	92
70) o-Xylene	12.06	106	5114	0.33	ug/l	95
85) 1,2,4-Trimethylbenzene	13.16	105	13944	0.44	ug/l	98
96) Naphthalene	15.29	128	351438	19.09	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA N\DATA\VN092115\
Data File : VN027362.D
Acq On : 22 Sep 2015 6:39
Operator : MD\FY
Sample : G3754-07
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-9B

Quant Time: Sep 22 08:22:33 2015
Quant Method : W:\HPCHEM1\MSVOA_N\METHODS\82N092115W.M
Quant Title : SW846 8260
QLast Update : Mon Sep 21 13:12:32 2015
Response via : Initial Calibration

