

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN092218\
 Data File : VN051378.D
 Acq On : 21 Sep 2018 19:02
 Operator : MD\SY
 Sample : PB113086ZHE#09
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 15 Sample Multiplier: 1


Quant Time: Sep 21 23:55:34 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N092018W.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 21 04:23:57 2018
 Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	756056	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1117337	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	948313	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	369642	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	421058	51.38	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.76%	
35) Dibromofluoromethane	7.59	113	420029	46.52	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.04%	
50) Toluene-d8	10.09	98	1496690	43.12	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.24%	
62) 4-Bromofluorobenzene	12.40	95	420399	37.23	ug/l	0.00
Spiked Amount 50.000			Recovery	=	74.46%	
Target Compounds						
16) Acetone	3.82	43	19455	10.712	ug/l	98
18) Methyl Acetate	4.33	43	11721	1.378	ug/l	99
20) Methylene Chloride	4.55	84	22804	2.707	ug/l	92
43) Isopropyl Acetate	8.17	43	286371	28.900	ug/l #	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN092218\
Data File : VN051378.D
Acq On : 21 Sep 2018 19:02
Operator : MD\SY
Sample : PB113086ZHE#09
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 15 Sample Multiplier: 1

Quant Time: Sep 21 23:55:34 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N092018W.M
Quant Title : SW846 8260
QLast Update : Fri Sep 21 04:23:57 2018
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.67 min Scan# 1834

Delta R.T. 0.00 min

Lab File: VN051378.D


Acq: 21 Sep 2018 19:02

Tot Ion:168 Resp: 756056

Ion Ratio Lower Upper

168 100

99 46.3 37.1 55.7



Abundance Ion 167.90 (167.60 to 168.60): VN051339.D

Ion 98.90 (98.60 to 99.60): VN051339.D

7.67

300000

200000

100000



0

Time-->

7.60

7.70

7.80



#16

Acetone

Concen: 10.712 ug/l

RT: 3.82 min Scan# 637

Delta R.T. 0.00 min

Lab File: VN051378.D


Acq: 21 Sep 2018 19:02

Tgt Ion: 43 Resp: 19455

Ion Ratio Lower Upper

43 100

58 36.7 28.3 42.5



Abundance Ion 43.00 (42.70 to 43.70): VN051339.D

Ion 58.00 (57.70 to 58.70): VN051339.D

3.82

8000

6000

4000

2000

0


Time-->


3.75

3.80

3.85





3.90








#18
Methvl Acetate
Concen: 1.378 ug/l
RT: 4.33 min Scan# 796
Delta R.T. 0.00 min
Lab File: VN051378.D
Acq: 21 Sep 2018 19:02


Tot Ion: 43 Resp: 11721
Ion Ratio Lower Upper
43 100
74 27.4 21.7 32.5



#20
Methylene Chloride
Concen: 2.707 ug/l
RT: 4.55 min Scan# 864
Delta R.T. -0.00 min
Lab File: VN051378.D
Acq: 21 Sep 2018 19:02

Tgt Ion: 84 Resp: 22804
Ion Ratio Lower Upper
84 100
49 114.5 83.2 124.8
51 33.3 26.4 39.6
86 71.5 51.4 77.0





#33

1,2-Dichloroethane-d4

Concen: 51.383 ug/l

RT: 8.03 min Scan# 1945

Delta R.T. -0.00 min

Lab File: VN051378.D


Acq: 21 Sep 2018 19:02

Tot Ion: 65 Resp: 421058

Ion Ratio Lower Upper

65 100

67 54.7 0.0 108.8



Abundance Ion 64.90 (64.60 to 65.60): VN0

Ion 66.90 (66.60 to 67.60): VN0

8.03



150000

100000

50000

0

Time--> 7.90 8.00 8.10



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.59 min Scan# 2119

Delta R.T. 0.00 min

Lab File: VN051378.D

Acq: 21 Sep 2018 19:02


Tgt Ion: 114 Resp: 1117337

Ion Ratio Lower Upper

114 100

63 17.9 0.0 36.6

88 15.1 0.0 29.8



Abundance Ion 113.90 (113.60 to 114.60): V

Ion 63.00 (62.70 to 63.70): VN0

Ion 87.90 (87.60 to 88.60): VN0


600000


400000

200000

0





Time--> 8.50 8.60 8.70








#35
 Dibromofluoromethane
 Concen: 46.517 ug/l
 RT: 7.59 min Scan# 1810
 Delta R.T. -0.00 min
 Lab File: VN051378.D
 Acq: 21 Sep 2018 19:02


Tot Ion	Ratio	Lower	Upper
113	100		
111	101.7	81.9	122.9
192	24.6	19.8	29.6



#43
 Isopropyl Acetate
 Concen: 28.900 ug/l
 RT: 8.17 min Scan# 1989
 Delta R.T. 0.00 min
 Lab File: VN051378.D
 Acq: 21 Sep 2018 19:02





Tgt Ion	Ratio	Lower	Upper
43	100		
61	16.3	18.1	27.1#
87	11.5	12.8	19.2#








#50
Toluene-d8
Concen: 43.117 ug/l
RT: 10.09 min Scan# 2587
Delta R.T. -0.00 min
Lab File: VN051378.D
Acq: 21 Sep 2018 19:02


Tot Ion: 98 Resp: 1496690
Ion Ratio Lower Upper
98 100
100 64.1 51.3 76.9



#62
4-Bromofluorobenzene
Concen: 37.227 ug/l
RT: 12.40 min Scan# 3306
Delta R.T. 0.00 min
Lab File: VN051378.D
Acq: 21 Sep 2018 19:02


Tgt Ion: 95 Resp: 420399
Ion Ratio Lower Upper
95 100
174 98.3 0.0 194.6
176 94.8 0.0 189.6








#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.41 min Scan# 2997
Delta R.T. 0.00 min
Lab File: VN051378.D
Acq: 21 Sep 2018 19:02

Tot Ion	Ratio	Lower	Upper
117	100		
82	51.3	40.7	61.1
119	32.6	25.3	37.9




Abundance Ion 116.90 (116.60 to 117.60): V
Ion 82.00 (81.70 to 82.70): V
Ion 118.90 (118.60 to 119.60): V



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.35 min Scan# 3599
Delta R.T. 0.00 min
Lab File: VN051378.D
Acq: 21 Sep 2018 19:02

Tgt Ion	Ratio	Lower	Upper
152	100		
115	53.9	26.9	80.7
150	156.5	0.0	345.0



Abundance Ion 151.90 (151.60 to 152.60): V
Ion 114.90 (114.60 to 115.60): V
Ion 149.90 (149.60 to 150.60): V

