

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN092218\
 Data File : VN051383.D
 Acq On : 21 Sep 2018 21:05
 Operator : MD\SY
 Sample : PB113086ZHE#14
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 20 Sample Multiplier: 1


Quant Time: Sep 22 00:15:35 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N092018W.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 21 04:23:57 2018
 Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	748171	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1115428	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	936537	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	355035	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	417043	51.43	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.86%	
35) Dibromofluoromethane	7.59	113	410435	45.53	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.06%	
50) Toluene-d8	10.09	98	1498168	43.23	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.46%	
62) 4-Bromofluorobenzene	12.40	95	408691	36.25	ug/l	0.00
Spiked Amount 50.000			Recovery	=	72.50%	
Target Compounds						
16) Acetone	3.82	43	18875	10.450	ug/l	100
18) Methyl Acetate	4.33	43	12928	1.662	ug/l #	73
20) Methylene Chloride	4.55	84	22440	2.692	ug/l	95
43) Isopropyl Acetate	8.17	43	270182	27.313	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN092218\
Data File : VN051383.D
Acq On : 21 Sep 2018 21:05
Operator : MD\SY
Sample : PB113086ZHE#14
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 20 Sample Multiplier: 1

Quant Time: Sep 22 00:15:35 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N092018W.M
Quant Title : SW846 8260
QLast Update : Fri Sep 21 04:23:57 2018
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.67 min Scan# 1833

Delta R.T. 0.00 min

Lab File: VN051383.D


Acq: 21 Sep 2018 21:05

Tot Ion:168 Resp: 748171

Ion Ratio Lower Upper

168 100

99 46.6 37.1 55.7



Abundance Ion 167.90 (167.60 to 168.60): VN051383.D (-1740) (-)

Ion 98.90 (98.60 to 99.60): VN051383.D (-1740) (-)

7.67

300000



200000

100000

0

7.60 7.70 7.80

Time-->



#16

Acetone

Concen: 10.450 ug/l

RT: 3.82 min Scan# 637

Delta R.T. 0.00 min

Lab File: VN051383.D


Acq: 21 Sep 2018 21:05

Tgt Ion: 43 Resp: 18875

Ion Ratio Lower Upper

43 100

58 35.4 28.3 42.5



Abundance Ion 43.00 (42.70 to 43.70): VN051383.D (-544) (-)

Ion 58.00 (57.70 to 58.70): VN051383.D (-544) (-)

3.82

8000

6000


4000


2000

0

3.70 3.80 3.90





Time-->








#18
Methvl Acetate
Concen: 1.662 ug/l
RT: 4.33 min Scan# 795
Delta R.T. -0.00 min
Lab File: VN051383.D
Acq: 21 Sep 2018 21:05


Tot Ion: 43 Resp: 12928
Ion Ratio Lower Upper
43 100
74 13.0 21.7 32.5#



#20
Methylene Chloride
Concen: 2.692 ug/l
RT: 4.55 min Scan# 865
Delta R.T. 0.00 min
Lab File: VN051383.D
Acq: 21 Sep 2018 21:05

Tgt Ion: 84 Resp: 22440
Ion Ratio Lower Upper
84 100
49 97.3 83.2 124.8
51 36.5 26.4 39.6
86 65.6 51.4 77.0





#33

1,2-Dichloroethane-d4

Concen: 51.429 ug/l

RT: 8.03 min Scan# 1946

Delta R.T. 0.00 min

Lab File: VN051383.D


Acq: 21 Sep 2018 21:05

Tot Ion: 65 Resp: 417043

Ion Ratio Lower Upper




65 100

67 54.4 0.0 108.8



Abundance Ion 64.90 (64.60 to 65.60): VN0

Ion 66.90 (66.60 to 67.60): VN0



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.59 min Scan# 2119

Delta R.T. 0.00 min

Lab File: VN051383.D

Acq: 21 Sep 2018 21:05


Tgt Ion: 114 Resp: 1115428

Ion Ratio Lower Upper

114 100

63 17.8 0.0 36.6



88 15.0 0.0 29.8




Abundance Ion 113.90 (113.60 to 114.60): V

Ion 63.00 (62.70 to 63.70): VN0





Ion 87.90 (87.60 to 88.60): VN0








#35
 Dibromofluoromethane
 Concen: 45.532 ug/l
 RT: 7.59 min Scan# 1810
 Delta R.T. 0.00 min
 Lab File: VN051383.D
 Acq: 21 Sep 2018 21:05


Tot Ion	Ratio	Lower	Upper
113	100		
111	104.3	81.9	122.9
192	25.0	19.8	29.6



#43
 Isopropyl Acetate
 Concen: 27.313 ug/l
 RT: 8.17 min Scan# 1989
 Delta R.T. 0.00 min
 Lab File: VN051383.D
 Acq: 21 Sep 2018 21:05





Tgt Ion	Ratio	Lower	Upper
43	100		
61	17.6	18.1	27.1#
87	11.9	12.8	19.2#








#50
Toluene-d8
Concen: 43.234 ug/l
RT: 10.09 min Scan# 2587
Delta R.T. 0.00 min
Lab File: VN051383.D
Acq: 21 Sep 2018 21:05


Tot Ion: 98 Resp: 1498168
Ion Ratio Lower Upper
98 100
100 63.8 51.3 76.9



#62
4-Bromofluorobenzene
Concen: 36.252 ug/l
RT: 12.40 min Scan# 3306
Delta R.T. 0.00 min
Lab File: VN051383.D
Acq: 21 Sep 2018 21:05


Tgt Ion: 95 Resp: 408691
Ion Ratio Lower Upper
95 100
174 98.2 0.0 194.6
176 95.9 0.0 189.6








#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.41 min Scan# 2996
Delta R.T. 0.00 min
Lab File: VN051383.D
Acq: 21 Sep 2018 21:05

Tot Ion:117 Resp: 936537
Ion Ratio Lower Upper
117 100
82 51.3 40.7 61.1
119 32.4 25.3 37.9




Abundance Ion 116.90 (116.60 to 117.60): V
Ion 82.00 (81.70 to 82.70): VN
Ion 118.90 (118.60 to 119.60): V



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.35 min Scan# 3599
Delta R.T. 0.00 min
Lab File: VN051383.D
Acq: 21 Sep 2018 21:05

Tgt Ion:152 Resp: 355035
Ion Ratio Lower Upper
152 100
115 54.6 26.9 80.7
150 156.6 0.0 345.0



Abundance Ion 151.90 (151.60 to 152.60): V
Ion 114.90 (114.60 to 115.60): V
Ion 149.90 (149.60 to 150.60): V

