

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN092718\
 Data File : VN051567.D
 Acq On : 27 Sep 2018 8:08
 Operator : MD\SY
 Sample : J4775-12
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 OR-01-092418-B

Quant Time: Sep 27 09:12:48 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N092318W.M
 Quant Title : SW846 8260
 QLast Update : Mon Sep 24 06:41:32 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	670767	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	1054318	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	892865	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	305315	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	464155	55.24	ug/l	0.00
Spiked Amount			Recovery	=	110.48%	
35) Dibromofluoromethane	7.59	113	416173	49.36	ug/l	0.00
Spiked Amount			Recovery	=	98.72%	
50) Toluene-d8	10.09	98	1496871	47.15	ug/l	0.00
Spiked Amount			Recovery	=	94.30%	
62) 4-Bromofluorobenzene	12.40	95	393076	38.27	ug/l	0.00
Spiked Amount			Recovery	=	76.54%	


Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
5) Bromomethane	2.57	94	1203	0.242	ug/l #	51
10) Methyl Iodide	3.96	142	615	0.333	ug/l #	42
11) Tert butyl alcohol	4.80	59	237	0.500	ug/l #	78
16) Acetone	3.82	43	36919	Below Cal		99
18) Methyl Acetate	4.33	43	9990	1.561	ug/l #	88
20) Methylene Chloride	4.54	84	16789	2.019	ug/l	98
25) 2-Butanone	6.85	43	2585	0.899	ug/l	91
29) Tetrahydrofuran	7.22	42	1040	0.571	ug/l #	49
30) Chloroform	7.37	83	10153	0.710	ug/l	88
31) Cyclohexane	7.66	56	13486	1.019	ug/l #	1
36) 1,1-Dichloropropene	7.66	75	46942	3.845	ug/l #	48
38) Carbon Tetrachloride	7.66	117	59527	4.661	ug/l #	16
43) Isopropyl Acetate	8.16	43	230156	18.131	ug/l	97
47) Bromodichloromethane	9.40	83	3440	0.282	ug/l #	87
48) Methyl methacrylate	9.18	41	6126	0.968	ug/l #	22
51) 4-Methyl-2-Pentanone	9.99	43	2680	0.384	ug/l #	87
59) 2-Hexanone	10.77	43	15629	3.453	ug/l #	56
71) Bromoform	12.40	173	4673	0.789	ug/l #	1
76) 1,2,3-Trichloropropane	12.40	75	194109	38.298	ug/l #	100
81) trans-1,4-Dichloro-2-buten	12.40	75	194109	117.763	ug/l #	12
93) 1,2,4-Trichlorobenzene	14.91	180	1452	6.198	ug/l	87
95) Naphthalene	15.14	128	6570	8.530	ug/l	98
96) 1,2,3-Trichlorobenzene	15.32	180	1386	0.264	ug/l	85


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN092718\
Data File : VN051567.D
Acq On : 27 Sep 2018 8:08
Operator : MD\SY
Sample : J4775-12
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 32 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
OR-01-092418-B

Quant Time: Sep 27 09:12:48 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N092318W.M
Quant Title : SW846 8260
QLast Update : Mon Sep 24 06:41:32 2018
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.66 min Scan# 1832

Delta R.T. -0.00 min

Lab File: VN051567.D

Acq: 27 Sep 2018 8:08

Instrument :

MSVOA_N

ClientSampleId :


OR-01-092418-B

Tot Ion:168 Resp: 670767

Ion Ratio Lower Upper

168 100

99 51.5 40.2 60.2



Abundance Ion 167.90 (167.60 to 168.60): VN051393.D (-1813) (-)

Ion 98.90 (98.60 to 99.60): VN051393.D (-1813) (-)

7.66

250000

200000



150000

100000

50000

0

Time-->



#5

Bromomethane

Concen: 0.242 ug/l

RT: 2.57 min Scan# 247

Delta R.T. 0.00 min

Lab File: VN051567.D


Acq: 27 Sep 2018 8:08

Tgt Ion: 94 Resp: 1203

Ion Ratio Lower Upper

94 100

96 44.9 73.6 110.4#



Abundance Ion 93.90 (93.60 to 94.60): VN051393.D (-228) (-)

Ion 95.90 (95.60 to 96.60): VN051393.D (-228) (-)

2.57

1000

800


600


400

200

0

Time-->





#10

Methvl Iodide

Concen: 0.333 ug/l

RT: 3.96 min Scan# 679

Delta R.T. 0.00 min

Lab File: VN051567.D

Acq: 27 Sep 2018 8:08

Instrument :

MSVOA_N

ClientSampleId :

OR-01-092418-B


Tot Ion:142 Resp: 615

Ion Ratio Lower Upper

142 100

127 0.0 33.2 49.8#




141 0.0 11.5 17.3#



Abundance Ion 141.80 (141.50 to 142.50): V

Ion 126.80 (126.50 to 127.50): V

Ion 140.80 (140.50 to 141.50): V



#11

Tert butyl alcohol

Concen: 0.500 ug/l

RT: 4.80 min Scan# 941

Delta R.T. 0.01 min

Lab File: VN051567.D


Acq: 27 Sep 2018 8:08

Tgt Ion: 59 Resp: 237

Ion Ratio Lower Upper



59 100


57 0.0 6.2 9.2#



Abundance Ion 59.00 (58.70 to 59.70): V

Ion 57.00 (56.70 to 57.70): V





#16

Acetone

Concen: Below Cal

RT: 3.82 min Scan# 635

Delta R.T. -0.00 min

Lab File: VN051567.D

Acq: 27 Sep 2018 8:08

Instrument :

MSVOA_N

ClientSampleId :


OR-01-092418-B

Tot Ion: 43 Resp: 36919

Ion Ratio Lower Upper




43 100

58 31.7 25.0 37.4



Abundance Ion 43.00 (42.70 to 43.70): VN0

Ion 58.00 (57.70 to 58.70): VN0



#18

Methyl Acetate

Concen: 1.561 ug/l

RT: 4.33 min Scan# 795

Delta R.T. 0.00 min

Lab File: VN051567.D


Acq: 27 Sep 2018 8:08

Tgt Ion: 43 Resp: 9990

Ion Ratio Lower Upper



43 100


74 17.7 18.8 28.2#



Abundance Ion 43.00 (42.70 to 43.70): VN0

Ion 74.00 (73.70 to 74.70): VN0









#20
Methvlene Chloride
Concen: 2.019 ug/l
RT: 4.54 min Scan# 861
Delta R.T. -0.01 min
Lab File: VN051567.D
Acq: 27 Sep 2018 8:08




Instrument :
MSVOA_N
ClientSampleId :
OR-01-092418-B


Tot Ion: 84 Resp: 16789
Ion Ratio Lower Upper
84 100
49 125.1 99.9 149.9
51 33.2 31.4 47.2
86 64.8 50.9 76.3



#25
2-Butanone
Concen: 0.899 ug/l
RT: 6.85 min Scan# 1578
Delta R.T. 0.01 min
Lab File: VN051567.D
Acq: 27 Sep 2018 8:08

Tgt Ion: 43 Resp: 2585
Ion Ratio Lower Upper
43 100
72 22.7 21.8 32.6





#29

Tetrahydrofuran

Concen: 0.571 ug/l

RT: 7.22 min Scan# 1693

Delta R.T. 0.00 min

Lab File: VN051567.D

Acq: 27 Sep 2018 8:08

Instrument :

MSVOA_N

ClientSampled :

OR-01-092418-B


Tot Ion: 42 Resp: 1040

Ion Ratio Lower Upper

42 100

72 18.5 34.3 51.5#


71 0.0 32.2 48.4#





Abundance Ion 42.00 (41.70 to 42.70): VN0

Ion 72.00 (71.70 to 72.70): VN0

Ion 71.00 (70.70 to 71.70): VN0



Time-->



#30

Chloroform

Concen: 0.710 ug/l

RT: 7.37 min Scan# 1741

Delta R.T. -0.00 min

Lab File: VN051567.D


Acq: 27 Sep 2018 8:08

Tgt Ion: 83 Resp: 10153

Ion Ratio Lower Upper


83 100

85 74.5 51.8 77.6





Abundance Ion 82.90 (82.60 to 83.60): VN0

Ion 84.90 (84.60 to 85.60): VN0



Time-->





#31

Cyclohexane

Concen: 1.019 ug/l

RT: 7.66 min Scan# 1832

Delta R.T. 0.01 min

Lab File: VN051567.D

Acq: 27 Sep 2018 8:08

Instrument :

MSVOA_N

ClientSampleId :

OR-01-092418-B


Tot Ion: 56 Resp: 13486

Ion Ratio Lower Upper

56 100

69 182.4 25.8 38.6#




84 146.2 66.0 99.0#



Abundance Ion 56.00 (55.70 to 56.70): VN051567.D (-1737) (-)

Ion 69.00 (68.70 to 69.70): VN051567.D (-1737) (-)

Ion 84.00 (83.70 to 84.70): VN051567.D (-1737) (-)



#33

1,2-Dichloroethane-d4

Concen: 55.240 ug/l

RT: 8.03 min Scan# 1945

Delta R.T. 0.00 min

Lab File: VN051567.D


Acq: 27 Sep 2018 8:08

Tgt Ion: 65 Resp: 464155

Ion Ratio Lower Upper



65 100


67 53.9 0.0 107.8



Abundance Ion 64.90 (64.60 to 65.60): VN051567.D (-1852) (-)

Ion 66.90 (66.60 to 67.60): VN051567.D (-1852) (-)









#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 8.58 min Scan# 2118
 Delta R.T. -0.00 min
 Lab File: VN051567.D
 Acq: 27 Sep 2018 8:08

Instrument :
 MSVOA_N
 ClientSampleId :
 OR-01-092418-B

Tot Ion:114 Resp: 1054318
 Ion Ratio Lower Upper
 114 100
 63 20.5 0.0 40.2
 88 15.4 0.0 31.4




Abundance Ion 113.90 (113.60 to 114.60): V
 Ion 63.00 (62.70 to 63.70): VN
 Ion 87.90 (87.60 to 88.60): VN






#35
 Dibromofluoromethane
 Concen: 49.363 ug/l
 RT: 7.59 min Scan# 1809
 Delta R.T. -0.00 min
 Lab File: VN051567.D
 Acq: 27 Sep 2018 8:08

Tgt Ion:113 Resp: 416173
 Ion Ratio Lower Upper
 113 100
 111 102.0 82.6 123.8
 192 21.7 18.0 27.0



Abundance Ion 112.80 (112.50 to 113.50): V
 Ion 110.80 (110.50 to 111.50): V
 Ion 191.70 (191.40 to 192.40): V





#36

1,1-Dichloropropene

Concen: 3.845 ug/l

RT: 7.66 min Scan# 1832

Delta R.T. -0.13 min

Lab File: VN051567.D

Acq: 27 Sep 2018 8:08

Instrument :

MSVOA_N

ClientSampleId :

OR-01-092418-B





Tot Ion: 75 Resp: 46942

Ion Ratio Lower Upper

75 100

110 8.6 18.1 54.3#

77 0.1 25.6 38.4#



#38

Carbon Tetrachloride

Concen: 4.661 ug/l

RT: 7.66 min Scan# 1832

Delta R.T. -0.11 min

Lab File: VN051567.D

Acq: 27 Sep 2018 8:08




Tgt Ion: 117 Resp: 59527


Ion Ratio Lower Upper

117 100

119 5.5 76.9 115.3#

121 0.0 24.6 37.0#






#43



Isopropyl Acetate
 Concen: 18.131 ug/l
 RT: 8.16 min Scan# 1987
 Delta R.T. -0.00 min
 Lab File: VN051567.D
 Acq: 27 Sep 2018 8:08

Instrument :
 MSVOA_N
 ClientSampleId :
 OR-01-092418-B


Tot Ion	Ratio	Lower	Upper
43	100		
61	19.1	16.8	25.2
87	12.4	10.7	16.1



Abundance Ion 43.00 (42.70 to 43.70): VNO
 Ion 61.00 (60.70 to 61.70): VNO
 Ion 87.00 (86.70 to 87.70): VNO




Time-->





#47


Bromodichloromethane
 Concen: 0.282 ug/l
 RT: 9.40 min Scan# 2372
 Delta R.T. -0.00 min
 Lab File: VN051567.D
 Acq: 27 Sep 2018 8:08

Tgt Ion	Ratio	Lower	Upper
83	100		
85	73.1	51.8	77.6
127	0.0	7.4	11.2#



Abundance Ion 82.90 (82.60 to 83.60): VNO
 Ion 84.90 (84.60 to 85.60): VNO
 Ion 126.80 (126.50 to 127.50): V









#48
Methyl methacrylate
Concen: 0.968 ug/l
RT: 9.18 min Scan# 2302
Delta R.T. -0.03 min
Lab File: VN051567.D
Acq: 27 Sep 2018 8:08

Instrument :
MSVOA_N
ClientSampleId :
OR-01-092418-B

Tot Ion: 41 Resp: 6126
Ion Ratio Lower Upper
41 100
69 0.0 70.4 105.6#
39 18.5 43.8 65.8#




Abundance Ion 41.00 (40.70 to 41.70): VN0
Ion 69.00 (68.70 to 69.70): VN0
Ion 39.00 (38.70 to 39.70): VN0






#50
Toluene-d8
Concen: 47.149 ug/l
RT: 10.09 min Scan# 2587
Delta R.T. 0.00 min
Lab File: VN051567.D
Acq: 27 Sep 2018 8:08

Tgt Ion: 98 Resp: 1496871
Ion Ratio Lower Upper
98 100
100 62.7 50.6 75.8



Abundance Ion 98.00 (97.70 to 98.70): VN0
Ion 100.00 (99.70 to 100.70): VN0









#51
4-Methyl-2-Pentanone
Concen: 0.384 ug/l
RT: 9.99 min Scan# 2554
Delta R.T. 0.00 min
Lab File: VN051567.D
Acq: 27 Sep 2018 8:08




Instrument :
MSVOA_N
ClientSampleId :
OR-01-092418-B


Tot Ion: 43 Resp: 2680
Ion Ratio Lower Upper
43 100
58 31.5 31.5 47.3#



#59
2-Hexanone
Concen: 3.453 ug/l
RT: 10.77 min Scan# 2797
Delta R.T. 0.01 min
Lab File: VN051567.D
Acq: 27 Sep 2018 8:08

Tgt Ion: 43 Resp: 15629
Ion Ratio Lower Upper
43 100
58 23.1 27.7 83.0#






#62
4-Bromofluorobenzene
Concen: 38.273 ug/l
RT: 12.40 min Scan# 3305
Delta R.T. 0.00 min
Lab File: VN051567.D
Acq: 27 Sep 2018 8:08

Instrument :
MSVOA_N
ClientSampleId :
OR-01-092418-B




Tot Ion	Ratio	Lower	Upper
95	100		
174	89.9	0.0	181.6
176	86.1	0.0	176.4



Abundance Ion 94.90 (94.60 to 95.60): VN051393.D (-3290) (-)


Ion 173.80 (173.50 to 174.50): VN051393.D (-3290) (-)

Ion 175.80 (175.50 to 176.50): VN051393.D (-3290) (-)



#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.41 min Scan# 2996
Delta R.T. 0.00 min
Lab File: VN051567.D
Acq: 27 Sep 2018 8:08



Tgt Ion	Ratio	Lower	Upper
117	100		
82	55.5	44.2	66.2
119	32.2	25.9	38.9




Abundance Ion 116.90 (116.60 to 117.60): VN051393.D (-2981) (-)

Ion 82.00 (81.70 to 82.70): VN051393.D (-2981) (-)

Ion 118.90 (118.60 to 119.60): VN051393.D (-2981) (-)









#71
Bromoform
Concen: 0.789 ug/l
RT: 12.40 min Scan# 3305
Delta R.T. 0.27 min
Lab File: VN051567.D
Acq: 27 Sep 2018 8:08

Instrument :
MSVOA_N
ClientSampleId :
OR-01-092418-B

Tot Ion:173 Resp: 4673
Ion Ratio Lower Upper
173 100
175 493.6 24.8 74.3#
254 0.0 0.2 0.2#




Abundance Ion 172.70 (172.40 to 173.40): V
Ion 174.70 (174.40 to 175.40): V
Ion 254.40 (254.10 to 255.10): V






#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.34 min Scan# 3598
Delta R.T. 0.00 min
Lab File: VN051567.D
Acq: 27 Sep 2018 8:08

Tgt Ion:152 Resp: 305315
Ion Ratio Lower Upper
152 100
115 58.0 28.3 84.9
150 159.4 0.0 353.6



Abundance Ion 151.90 (151.60 to 152.60): V
Ion 114.90 (114.60 to 115.60): V
Ion 149.90 (149.60 to 150.60): V





#76

1,2,3-Trichloropropane

Concen: 38.298 ug/l

RT: 12.40 min Scan# 3305

Delta R.T. -0.15 min

Lab File: VN051567.D

Acq: 27 Sep 2018 8:08

Instrument :

MSVOA_N

ClientSampleId :


OR-01-092418-B

Tot Ion: 75 Resp: 194109

Ion Ratio Lower Upper




75 100

77 1.2 0.0 0.0#



Abundance Ion 74.90 (74.60 to 75.60): VNO

Ion 77.00 (76.70 to 77.70): VNO



#81

trans-1,4-Dichloro-2-butene

Concen: 117.763 ug/l

RT: 12.40 min Scan# 3305

Delta R.T. 0.10 min

Lab File: VN051567.D

Acq: 27 Sep 2018 8:08


Tgt Ion: 75 Resp: 194109

Ion Ratio Lower Upper

75 100



53 0.2 73.9 110.9#


89 0.0 36.7 55.1#



Abundance Ion 74.90 (74.60 to 75.60): VNO

Ion 53.00 (52.70 to 53.70): VNO









#93
1,2,4-Trichlorobenzene
Concen: 6.198 ug/l
RT: 14.91 min Scan# 4086
Delta R.T. 0.00 min
Lab File: VN051567.D
Acq: 27 Sep 2018 8:08

Instrument :
MSVOA_N
ClientSampleId :
OR-01-092418-B

Tot Ion:180 Resp: 1452
Ion Ratio Lower Upper
180 100
182 84.6 48.5 145.6
145 20.5 14.1 42.4




Abundance Ion 179.80 (179.50 to 180.50): V
Ion 181.80 (181.50 to 182.50): V
Ion 144.80 (144.50 to 145.50): V






#95
Naphthalene
Concen: 8.530 ug/l
RT: 15.14 min Scan# 4156
Delta R.T. 0.00 min
Lab File: VN051567.D
Acq: 27 Sep 2018 8:08

Tgt Ion:128 Resp: 6570
Ion Ratio Lower Upper
128 100
127 11.9 10.5 15.7
129 10.4 8.6 12.8



Abundance Ion 128.00 (127.70 to 128.70): V
Ion 127.00 (126.70 to 127.70): V
Ion 129.00 (128.70 to 129.70): V





#96
 1,2,3-Trichlorobenzene
 Concen: 0.264 ug/l
 RT: 15.32 min Scan# 4212
 Delta R.T. 0.00 min
 Lab File: VN051567.D
 Acq: 27 Sep 2018 8:08

Instrument :
 MSVOA_N
 ClientSampleId :
 OR-01-092418-B

Tot Ion	Ratio	Lower	Upper
180	100		
182	108.3	47.2	141.6
145	38.7	14.6	43.8

