

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN092718\
 Data File : VN051553.D
 Acq On : 26 Sep 2018 22:51
 Operator : MD\SY
 Sample : PB113219ZHE#10
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB113219ZHE#10

Quant Time: Sep 27 08:42:29 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N092318W.M
 Quant Title : SW846 8260
 QLast Update : Mon Sep 24 06:41:32 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	610741	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	999321	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	805999	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	257785	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	438307	57.29	ug/l	0.00
Spiked Amount 50.000			Recovery	=	114.58%	
35) Dibromofluoromethane	7.59	113	378514	47.37	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.74%	
50) Toluene-d8	10.09	98	1363922	45.33	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.66%	
62) 4-Bromofluorobenzene	12.40	95	342141	35.15	ug/l	0.00
Spiked Amount 50.000			Recovery	=	70.30%	
Target Compounds						
18) Methyl Acetate	4.34	43	17351	2.978	ug/l #	77
20) Methylene Chloride	4.55	84	9955	1.315	ug/l	96
43) Isopropyl Acetate	8.17	43	438639	36.457	ug/l #	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN092718\
Data File : VN051553.D
Acq On : 26 Sep 2018 22:51
Operator : MD\SY
Sample : PB113219ZHE#10
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB113219ZHE#10

Quant Time: Sep 27 08:42:29 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N092318W.M
Quant Title : SW846 8260
QLast Update : Mon Sep 24 06:41:32 2018
Response via : Initial Calibration

