

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN092718\
 Data File : VN051555.D
 Acq On : 26 Sep 2018 23:40
 Operator : MD\SY
 Sample : J5070-03
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 CS-RBR-200027

Quant Time: Sep 27 08:45:59 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N092318W.M
 Quant Title : SW846 8260
 QLast Update : Mon Sep 24 06:41:32 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	609105	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1003466	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	858013	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	338887	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	452631	59.32	ug/l	0.00
Spiked Amount						
			Recovery	=	118.64%	
35) Dibromofluoromethane	7.60	113	387515	48.29	ug/l	0.00
Spiked Amount						
			Recovery	=	96.58%	
50) Toluene-d8	10.09	98	1396371	46.21	ug/l	0.00
Spiked Amount						
			Recovery	=	92.42%	
62) 4-Bromofluorobenzene	12.40	95	399812	40.90	ug/l	0.00
Spiked Amount						
			Recovery	=	81.80%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
18) Methyl Acetate	4.34	43	19288	3.319	ug/l	98
20) Methylene Chloride	4.56	84	16000	2.119	ug/l	93
39) Methylcyclohexane	9.08	83	14863	1.094	ug/l	90
43) Isopropyl Acetate	8.17	43	343865	28.462	ug/l #	88
78) n-propylbenzene	12.59	91	42238	1.311	ug/l	98
95) Naphthalene	15.13	128	17273	9.195	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN092718\
Data File : VN051555.D
Acq On : 26 Sep 2018 23:40
Operator : MD\SY
Sample : J5070-03
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
CS-RBR-200027

Quant Time: Sep 27 08:45:59 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N092318W.M
Quant Title : SW846 8260
QLast Update : Mon Sep 24 06:41:32 2018
Response via : Initial Calibration

