

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN092718\
 Data File : VN051569.D
 Acq On : 27 Sep 2018 8:57
 Operator : MD\SY
 Sample : J4779-06
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SU-02-092418-A

Quant Time: Sep 27 11:31:29 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N092318W.M
 Quant Title : SW846 8260
 QLast Update : Mon Sep 24 06:41:32 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	650347	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1041135	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	862709	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	295148	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	459177	56.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	112.72%	
35) Dibromofluoromethane	7.59	113	400449	48.10	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.20%	
50) Toluene-d8	10.09	98	1464171	46.70	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.40%	
62) 4-Bromofluorobenzene	12.40	95	384757	37.94	ug/l	0.00
Spiked Amount 50.000			Recovery	=	75.88%	
Target Compounds						
20) Methylene Chloride	4.55	84	130238	16.155	ug/l	97
30) Chloroform	7.37	83	6814	0.491	ug/l	97
43) Isopropyl Acetate	8.17	43	230538	18.391	ug/l	94
95) Naphthalene	15.13	128	16446	9.302	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN092718\
Data File : VN051569.D
Acq On : 27 Sep 2018 8:57
Operator : MD\SY
Sample : J4779-06
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 34 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SU-02-092418-A

Quant Time: Sep 27 11:31:29 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N092318W.M
Quant Title : SW846 8260
QLast Update : Mon Sep 24 06:41:32 2018
Response via : Initial Calibration

