

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN092718\
 Data File : VN051570.D
 Acq On : 27 Sep 2018 9:22
 Operator : MD\SY
 Sample : J4779-08
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SU-02-092418-B

Quant Time: Sep 27 11:34:29 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N092318W.M
 Quant Title : SW846 8260
 QLast Update : Mon Sep 24 06:41:32 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	635713	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1025435	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	871415	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	321680	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	453420	56.94	ug/l	0.00
Spiked Amount			Recovery	=	113.88%	
35) Dibromofluoromethane	7.59	113	395715	48.26	ug/l	0.00
Spiked Amount			Recovery	=	96.52%	
50) Toluene-d8	10.09	98	1440691	46.66	ug/l	0.00
Spiked Amount			Recovery	=	93.32%	
62) 4-Bromofluorobenzene	12.40	95	399129	39.96	ug/l	0.00
Spiked Amount			Recovery	=	79.92%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
20) Methylene Chloride	4.55	84	54805	6.955	ug/l	97
30) Chloroform	7.37	83	9676	0.713	ug/l	98
43) Isopropyl Acetate	8.17	43	261466	21.178	ug/l	94
67) Ethyl Benzene	11.51	91	11206	0.297	ug/l	99
68) m/p-Xylenes	11.62	106	12459	0.855	ug/l	95
69) o-Xylene	11.95	106	3738	0.267	ug/l	97
84) 1,2,4-Trimethylbenzene	13.04	105	10821	0.485	ug/l	98
95) Naphthalene	15.13	128	33888	10.422	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN092718\
Data File : VN051570.D
Acq On : 27 Sep 2018 9:22
Operator : MD\SY
Sample : J4779-08
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 35 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SU-02-092418-B

Quant Time: Sep 27 11:34:29 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N092318W.M
Quant Title : SW846 8260
QLast Update : Mon Sep 24 06:41:32 2018
Response via : Initial Calibration

