

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN092718\
 Data File : VN051580.D
 Acq On : 27 Sep 2018 13:28
 Operator : MD\SY
 Sample : J4773-14
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 45 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 CL-02-092518-B

Quant Time: Sep 27 14:38:50 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N092318W.M
 Quant Title : SW846 8260
 QLast Update : Mon Sep 24 06:41:32 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	632547	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1032947	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	872672	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	315276	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	456990	57.67	ug/l	0.00
Spiked Amount						
			Recovery	=	115.34%	
35) Dibromofluoromethane	7.59	113	399610	48.38	ug/l	0.00
Spiked Amount						
			Recovery	=	96.76%	
50) Toluene-d8	10.09	98	1430978	46.01	ug/l	0.00
Spiked Amount						
			Recovery	=	92.02%	
62) 4-Bromofluorobenzene	12.40	95	393808	39.14	ug/l	0.00
Spiked Amount						
			Recovery	=	78.28%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
20) Methylene Chloride	4.55	84	67651	8.628	ug/l	99
30) Chloroform	7.37	83	5491	0.407	ug/l	95
43) Isopropyl Acetate	8.17	43	321140	25.822	ug/l #	91
44) Trichloroethene	8.83	130	3442	0.369	ug/l	96
84) 1,2,4-Trimethylbenzene	13.04	105	6597	0.302	ug/l	96
95) Naphthalene	15.13	128	38609	10.808	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN092718\
Data File : VN051580.D
Acq On : 27 Sep 2018 13:28
Operator : MD\SY
Sample : J4773-14
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 45 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
CL-02-092518-B

Quant Time: Sep 27 14:38:50 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N092318W.M
Quant Title : SW846 8260
QLast Update : Mon Sep 24 06:41:32 2018
Response via : Initial Calibration

