

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN092718\
 Data File : VN051581.D
 Acq On : 27 Sep 2018 13:52
 Operator : MD\SY
 Sample : J4773-12
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 46 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 CL-02-092518-A

Quant Time: Sep 27 14:41:48 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N092318W.M
 Quant Title : SW846 8260
 QLast Update : Mon Sep 24 06:41:32 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	660446	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1070853	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	883919	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	317951	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	468602	56.64	ug/l	0.00
Spiked Amount 50.000			Recovery	=	113.28%	
35) Dibromofluoromethane	7.59	113	410373	47.92	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.84%	
50) Toluene-d8	10.09	98	1479599	45.89	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.78%	
62) 4-Bromofluorobenzene	12.40	95	397092	38.07	ug/l	0.00
Spiked Amount 50.000			Recovery	=	76.14%	
Target Compounds						
20) Methylene Chloride	4.55	84	387243	47.301	ug/l	98
30) Chloroform	7.37	83	7486	0.531	ug/l	96
43) Isopropyl Acetate	8.17	43	349359	27.097	ug/l #	88
95) Naphthalene	15.13	128	10584	8.796	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN092718\
Data File : VN051581.D
Acq On : 27 Sep 2018 13:52
Operator : MD\SY
Sample : J4773-12
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 46 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
CL-02-092518-A

Quant Time: Sep 27 14:41:48 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N092318W.M
Quant Title : SW846 8260
QLast Update : Mon Sep 24 06:41:32 2018
Response via : Initial Calibration

