

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN092718\
 Data File : VN051593.D
 Acq On : 27 Sep 2018 19:43
 Operator : MD\SY
 Sample : PB113263ZHE#05
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 58 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB113263ZHE#05

Quant Time: Sep 28 02:10:35 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N092318W.M
 Quant Title : SW846 8260
 QLast Update : Mon Sep 24 06:41:32 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	626114	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1038070	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	878601	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	317709	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	453503	57.82	ug/l	0.00
Spiked Amount	50.000		Recovery	=	115.64%	
35) Dibromofluoromethane	7.59	113	398309	47.98	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.96%	
50) Toluene-d8	10.09	98	1446286	46.27	ug/l	0.00
Spiked Amount	50.000		Recovery	=	92.54%	
62) 4-Bromofluorobenzene	12.40	95	393903	38.95	ug/l	0.00
Spiked Amount	50.000		Recovery	=	77.90%	
Target Compounds						
18) Methyl Acetate	4.34	43	14294	2.393	ug/l	95
20) Methylene Chloride	4.56	84	46836	6.035	ug/l	98
43) Isopropyl Acetate	8.17	43	342495	27.404	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN092718\
Data File : VN051593.D
Acq On : 27 Sep 2018 19:43
Operator : MD\SY
Sample : PB113263ZHE#05
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 58 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB113263ZHE#05

Quant Time: Sep 28 02:10:35 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N092318W.M
Quant Title : SW846 8260
QLast Update : Mon Sep 24 06:41:32 2018
Response via : Initial Calibration

