

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN092718\
 Data File : VN051594.D
 Acq On : 27 Sep 2018 20:08
 Operator : MD\SY
 Sample : PB113263ZHE#06
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 59 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB113263ZHE#06

Quant Time: Sep 28 02:12:59 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N092318W.M
 Quant Title : SW846 8260
 QLast Update : Mon Sep 24 06:41:32 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	617982	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1024427	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	842337	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	285940	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	453700	58.61	ug/l	0.00
Spiked Amount 50.000			Recovery	=	117.22%	
35) Dibromofluoromethane	7.59	113	395044	48.22	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.44%	
50) Toluene-d8	10.09	98	1418465	45.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.96%	
62) 4-Bromofluorobenzene	12.40	95	367680	36.84	ug/l	0.00
Spiked Amount 50.000			Recovery	=	73.68%	
Target Compounds						
18) Methyl Acetate	4.34	43	14304	2.426	ug/l	95
20) Methylene Chloride	4.56	84	46151	6.025	ug/l	98
43) Isopropyl Acetate	8.17	43	346993	28.133	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN092718\
Data File : VN051594.D
Acq On : 27 Sep 2018 20:08
Operator : MD\SY
Sample : PB113263ZHE#06
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 59 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB113263ZHE#06

Quant Time: Sep 28 02:12:59 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N092318W.M
Quant Title : SW846 8260
QLast Update : Mon Sep 24 06:41:32 2018
Response via : Initial Calibration

