

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN092718\
 Data File : VN051595.D
 Acq On : 27 Sep 2018 20:32
 Operator : MD\SY
 Sample : PB113263ZHE#07
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 60 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB113263ZHE#07

Quant Time: Sep 28 02:16:38 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N092318W.M
 Quant Title : SW846 8260
 QLast Update : Mon Sep 24 06:41:32 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	622591	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1048953	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	875931	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	297957	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	453715	58.18	ug/l	0.00
Spiked Amount 50.000			Recovery	=	116.36%	
35) Dibromofluoromethane	7.59	113	402069	47.93	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.86%	
50) Toluene-d8	10.09	98	1447944	45.84	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.68%	
62) 4-Bromofluorobenzene	12.40	95	384256	37.61	ug/l	0.00
Spiked Amount 50.000			Recovery	=	75.22%	
Target Compounds						
18) Methyl Acetate	4.34	43	18508	3.116	ug/l	95
20) Methylene Chloride	4.55	84	46512	6.027	ug/l	93
43) Isopropyl Acetate	8.17	43	366045	28.984	ug/l #	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN092718\
Data File : VN051595.D
Acq On : 27 Sep 2018 20:32
Operator : MD\SY
Sample : PB113263ZHE#07
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 60 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB113263ZHE#07

Quant Time: Sep 28 02:16:38 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N092318W.M
Quant Title : SW846 8260
QLast Update : Mon Sep 24 06:41:32 2018
Response via : Initial Calibration

