

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN092718\
 Data File : VN051597.D
 Acq On : 27 Sep 2018 21:22
 Operator : MD\SY
 Sample : J5081-16
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 62 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-2-D

Quant Time: Sep 28 02:25:07 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N092318W.M
 Quant Title : SW846 8260
 QLast Update : Mon Sep 24 06:41:32 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	621139	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1041511	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	857843	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	307323	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	459688	59.08	ug/l	0.00
Spiked Amount 50.000			Recovery	=	118.16%	
35) Dibromofluoromethane	7.59	113	399384	47.95	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.90%	
50) Toluene-d8	10.09	98	1427157	45.51	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.02%	
62) 4-Bromofluorobenzene	12.40	95	386183	38.06	ug/l	0.00
Spiked Amount 50.000			Recovery	=	76.12%	
Target Compounds						
18) Methyl Acetate	4.33	43	19389	3.272	ug/l	96
20) Methylene Chloride	4.55	84	150414	19.535	ug/l	98
43) Isopropyl Acetate	8.17	43	387758	30.923	ug/l #	85
95) Naphthalene	15.14	128	37156	10.773	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN092718\
Data File : VN051597.D
Acq On : 27 Sep 2018 21:22
Operator : MD\SY
Sample : J5081-16
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 62 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-2-D

Quant Time: Sep 28 02:25:07 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N092318W.M
Quant Title : SW846 8260
QLast Update : Mon Sep 24 06:41:32 2018
Response via : Initial Calibration

