


Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN092718\
 Data File : VN051538.D
 Acq On : 26 Sep 2018 15:31
 Operator : MD\SY
 Sample : BFB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N092318W.M
 Title : SW846 8260
 Last Update : Mon Sep 24 06:41:32 2018



AutoFind: Scans 3305, 3306, 3307; Background Corrected with Scan 3291

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	17.8	12280	PASS
75	95	30	60	47.4	32714	PASS
95	95	100	100	100.0	68952	PASS
96	95	5	9	7.1	4926	PASS
173	174	0.00	2	1.1	657	PASS
174	95	50	100	89.9	61994	PASS
175	174	5	9	7.0	4334	PASS
176	174	95	101	96.2	59634	PASS
177	176	5	9	6.7	3974	PASS