


Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN092718\
Data File : VN051562.D
Acq On : 27 Sep 2018 4:09
Operator : MD\SY
Sample : BFB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N092318W.M
Title : SW846 8260
Last Update : Mon Sep 24 06:41:32 2018



AutoFind: Scans 3305, 3306, 3307; Background Corrected with Scan 3292

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	17.5	8174	PASS
75	95	30	60	46.8	21840	PASS
95	95	100	100	100.0	46669	PASS
96	95	5	9	6.5	3047	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	88.8	41453	PASS
175	174	5	9	7.6	3169	PASS
176	174	95	101	97.0	40200	PASS
177	176	5	9	6.0	2429	PASS