

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101824\
 Data File : VN084423.D
 Acq On : 18 Oct 2024 12:38
 Operator : JC\MD
 Sample : P4427-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 241016065-06-TRIP-BLANK

Quant Time: Oct 19 03:50:21 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N100824W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Wed Oct 09 02:22:28 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	7.812	128	30955	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	9.100	114	148440	30.000	ug/l	0.00
57) Chlorobenzene-d5	11.865	117	135324	30.000	ug/l	0.00

System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.577	65	72386	29.045	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	96.833%
60) 4-Bromofluorobenzene	12.847	95	55514	24.856	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	82.867%
63) Toluene-d8	10.565	98	182678	29.139	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	97.133%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101824\
Data File : VN084423.D
Acq On : 18 Oct 2024 12:38
Operator : JC\MD
Sample : P4427-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
241016065-06-TRIP-BLANK

Quant Time: Oct 19 03:50:21 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N100824W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Wed Oct 09 02:22:28 2024
Response via : Initial Calibration

