



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102324\
Data File : VN084468.D
Acq On : 23 Oct 2024 12:39
Operator : JC\MD
Sample : BFB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BFB

Integration File: RTEINT3.P

Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N100824W.M
Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
Last Update : Wed Oct 09 02:22:28 2024



AutoFind: Scans 1851, 1852, 1853; Background Corrected with Scan 1845

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.4	6313	PASS
75	95	30	60	52.9	17204	PASS
95	95	100	100	100.0	32536	PASS
96	95	5	9	6.6	2161	PASS
173	174	0.00	2	1.1	270	PASS
174	95	50	100	74.2	24149	PASS
175	174	5	9	7.5	1817	PASS
176	174	95	101	94.7	22875	FAIL*
177	176	5	9	7.0	1598	PASS