

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102424\
 Data File : VN084478.D
 Acq On : 24 Oct 2024 10:42
 Operator : JC\MD
 Sample : VN1024WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1024WBL01

Quant Time: Oct 25 02:40:44 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 01 07:11:01 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	178165	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	302564	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	259401	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	101906	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	127572	48.293	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	96.580%
35) Dibromofluoromethane	8.165	113	104023	52.150	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	104.300%
50) Toluene-d8	10.565	98	364872	49.719	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.440%
62) 4-Bromofluorobenzene	12.847	95	113388	42.411	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	84.820%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102424\
Data File : VN084478.D
Acq On : 24 Oct 2024 10:42
Operator : JC\MD
Sample : VN1024WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1024WBL01

Quant Time: Oct 25 02:40:44 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 01 07:11:01 2024
Response via : Initial Calibration

