

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102424\
 Data File : VN084485.D
 Acq On : 24 Oct 2024 13:58
 Operator : JC\MD
 Sample : VIBLK
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VIBLK

Quant Time: Oct 25 02:00:14 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 01 07:11:01 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	194664	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	339701	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	305703	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	131696	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	131616	45.601	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	91.200%
35) Dibromofluoromethane	8.165	113	102909	45.951	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.900%
50) Toluene-d8	10.565	98	413938	50.239	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	100.480%
62) 4-Bromofluorobenzene	12.847	95	150602	50.172	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	100.340%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102424\
Data File : VN084485.D
Acq On : 24 Oct 2024 13:58
Operator : JC\MD
Sample : VIBLK
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VIBLK

Quant Time: Oct 25 02:00:14 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 01 07:11:01 2024
Response via : Initial Calibration

