

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102424\
 Data File : VN084493.D
 Acq On : 24 Oct 2024 17:09
 Operator : JC\MD
 Sample : PB164314ZHE#07
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB164314ZHE#07

Manual Integrations
 APPROVED

Reviewed By :Mahesh Dadoda 10/25/2024
 Supervised By :Semsettin Yesilyurt 10/25/2024

Quant Time: Oct 25 02:04:01 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 01 07:11:01 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	196176	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	361190	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	316668	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	129797	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.576	65	140010	48.135	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	96.280%		
35) Dibromofluoromethane	8.165	113	110586	46.441	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	92.880%		
50) Toluene-d8	10.565	98	432401	49.357	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	98.720%		
62) 4-Bromofluorobenzene	12.847	95	145784	45.678	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	91.360%		
Target Compounds					Qvalue	
16) Acetone	4.430	43	49850	39.896	ug/l	96
20) Methylene Chloride	5.271	84	5262m	2.072	ug/l	
43) Isopropyl Acetate	8.688	43	278855m	40.122	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102424\
Data File : VN084493.D
Acq On : 24 Oct 2024 17:09
Operator : JC\MD
Sample : PB164314ZHE#07
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB164314ZHE#07

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 10/25/2024
Supervised By :Semsettin Yesilyurt 10/25/2024

Quant Time: Oct 25 02:04:01 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 01 07:11:01 2024
Response via : Initial Calibration

