

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102424\
 Data File : VN084494.D
 Acq On : 24 Oct 2024 17:33
 Operator : JC\MD
 Sample : PB164314ZHE#08
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB164314ZHE#08

Manual Integrations
 APPROVED

Reviewed By :Mahesh Dadoda 10/25/2024
 Supervised By :Semsettin Yesilyurt 10/25/2024

Quant Time: Oct 25 02:04:32 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 01 07:11:01 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	193219	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	351046	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	309189	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	125619	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	138598	48.379	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	96.760%		
35) Dibromofluoromethane	8.165	113	109412	47.276	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	94.560%		
50) Toluene-d8	10.565	98	422847	49.661	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	99.320%		
62) 4-Bromofluorobenzene	12.847	95	143330	46.206	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	92.420%		
Target Compounds						
16) Acetone	4.436	43	29210	23.735	ug/l	91
20) Methylene Chloride	5.271	84	4699	1.878	ug/l #	72
43) Isopropyl Acetate	8.688	43	281298m	41.643	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102424\
Data File : VN084494.D
Acq On : 24 Oct 2024 17:33
Operator : JC\MD
Sample : PB164314ZHE#08
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB164314ZHE#08

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 10/25/2024
Supervised By :Semsettin Yesilyurt 10/25/2024

Quant Time: Oct 25 02:04:32 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 01 07:11:01 2024
Response via : Initial Calibration

