

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102524\
 Data File : VN084514.D
 Acq On : 25 Oct 2024 14:27
 Operator : JC\MD
 Sample : VIBLK
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VIBLK

Quant Time: Oct 26 05:29:47 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N100824W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Wed Oct 09 02:22:28 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	7.818	128	31809	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	9.100	114	158755	30.000	ug/l	0.00
57) Chlorobenzene-d5	11.865	117	141172	30.000	ug/l	0.00

System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.577	65	76383	29.826	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	99.433%
60) 4-Bromofluorobenzene	12.847	95	57373	24.624	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	82.067%
63) Toluene-d8	10.565	98	187562	28.679	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	95.600%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102524\
Data File : VN084514.D
Acq On : 25 Oct 2024 14:27
Operator : JC\MD
Sample : VIBLK
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VIBLK

Quant Time: Oct 26 05:29:47 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N100824W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Wed Oct 09 02:22:28 2024
Response via : Initial Calibration

