

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102524\
 Data File : VN084520.D
 Acq On : 25 Oct 2024 16:50
 Operator : JC\MD
 Sample : P4528-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 241023062-05-TRIP-BLANK

Quant Time: Oct 26 06:40:58 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 01 07:11:01 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	125033	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	236367	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	203498	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	80554	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	111771	60.291	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	120.580%
35) Dibromofluoromethane	8.165	113	90099	57.820	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	115.640%
50) Toluene-d8	10.565	98	287511	50.149	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	100.300%
62) 4-Bromofluorobenzene	12.847	95	95020	45.494	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	90.980%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102524\
Data File : VN084520.D
Acq On : 25 Oct 2024 16:50
Operator : JC\MD
Sample : P4528-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
241023062-05-TRIP-BLANK

Quant Time: Oct 26 06:40:58 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 01 07:11:01 2024
Response via : Initial Calibration

