

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102524\
 Data File : VN084525.D
 Acq On : 25 Oct 2024 18:49
 Operator : JC\MD
 Sample : P4550-05
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BP-VPB-190-GW-318-320

Quant Time: Oct 25 23:38:13 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 01 07:11:01 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	132840	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	246729	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	211932	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	84516	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	115498	58.640	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	117.280%
35) Dibromofluoromethane	8.165	113	90200	55.453	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	110.900%
50) Toluene-d8	10.565	98	287073	47.970	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	95.940%
62) 4-Bromofluorobenzene	12.847	95	91662	42.043	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	84.080%
Target Compounds						
					Qvalue	
16) Acetone	4.448	43	3457	4.086	ug/l	93
17) Carbon Disulfide	4.718	76	7509	1.586	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102524\
Data File : VN084525.D
Acq On : 25 Oct 2024 18:49
Operator : JC\MD
Sample : P4550-05
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-VPB-190-GW-318-320

Quant Time: Oct 25 23:38:13 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 01 07:11:01 2024
Response via : Initial Calibration

