

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102824\
 Data File : VN084538.D
 Acq On : 28 Oct 2024 14:41
 Operator : JC\MD
 Sample : P4550-12
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BP-VPB-190-GW-418-420

Manual Integrations
 APPROVED

Reviewed By :Mahesh Dadoda 10/29/2024
 Supervised By :Semsettin Yesilyurt 10/29/2024

Quant Time: Oct 28 18:34:36 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 01 07:11:01 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	158931	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	275224	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	239281	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	94178	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.576	65	115590	49.053	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	98.100%		
35) Dibromofluoromethane	8.165	113	93090	51.305	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	102.600%		
50) Toluene-d8	10.565	98	330849	49.561	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	99.120%		
62) 4-Bromofluorobenzene	12.847	95	110647	45.497	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	91.000%		
Target Compounds						
16) Acetone	4.436	43	4608m	4.552	ug/l	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102824\
Data File : VN084538.D
Acq On : 28 Oct 2024 14:41
Operator : JC\MD
Sample : P4550-12
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-VPB-190-GW-418-420

Quant Time: Oct 28 18:34:36 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 01 07:11:01 2024
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 10/29/2024
Supervised By :Semsettin Yesilyurt 10/29/2024

