

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102824\
 Data File : VN084547.D
 Acq On : 28 Oct 2024 18:17
 Operator : JC\MD
 Sample : P4545-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VNJ-215

Quant Time: Oct 29 01:32:27 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 01 07:11:01 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	158440	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	280484	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	259107	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	99602	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	115859	49.319	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	98.640%
35) Dibromofluoromethane	8.165	113	90938	49.179	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.360%
50) Toluene-d8	10.565	98	353124	51.906	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	103.820%
62) 4-Bromofluorobenzene	12.847	95	116528	47.017	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	94.040%
Target Compounds						
					Qvalue	
16) Acetone	4.436	43	38823	38.471	ug/l	97
20) Methylene Chloride	5.277	84	2813	1.371	ug/l	91
43) Isopropyl Acetate	8.688	43	87918	16.289	ug/l #	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102824\
Data File : VN084547.D
Acq On : 28 Oct 2024 18:17
Operator : JC\MD
Sample : P4545-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VNJ-215

Quant Time: Oct 29 01:32:27 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 01 07:11:01 2024
Response via : Initial Calibration

