

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102824\
 Data File : VN084534.D
 Acq On : 28 Oct 2024 13:06
 Operator : JC\MD
 Sample : VIBLK
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VIBLK

Quant Time: Oct 28 14:15:19 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 01 07:11:01 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	163289	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	283770	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	244760	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	100197	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	119009	49.156	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	98.320%
35) Dibromofluoromethane	8.165	113	95292	50.937	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.880%
50) Toluene-d8	10.565	98	340205	49.428	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.860%
62) 4-Bromofluorobenzene	12.847	95	116781	46.573	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	93.140%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102824\
Data File : VN084534.D
Acq On : 28 Oct 2024 13:06
Operator : JC\MD
Sample : VIBLK
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VIBLK

Quant Time: Oct 28 14:15:19 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 01 07:11:01 2024
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.224 min Scan# 1066
Delta R.T. -0.000 min
Lab File: VN084534.D
Acq: 28 Oct 2024 13:06




Instrument :
MSVOA_N
ClientSampleId :
VIBLK


Tgt Ion:168 Resp: 163289
Ion Ratio Lower Upper
168 100
99 63.1 54.2 81.2



#33
1,2-Dichloroethane-d4
Concen: 49.156 ug/l
RT: 8.577 min Scan# 1126
Delta R.T. -0.000 min
Lab File: VN084534.D
Acq: 28 Oct 2024 13:06

Tgt Ion: 65 Resp: 119009
Ion Ratio Lower Upper
65 100
67 51.5 0.0 102.0





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.100 min Scan# 1

Delta R.T. -0.000 min

Lab File: VN084534.D

Acq: 28 Oct 2024 13:06

Instrument :

MSVOA_N

ClientSampleId :

VIBLK





Tgt Ion:114 Resp: 283770

Ion Ratio Lower Upper

114 100

63 21.1 0.0 43.8

88 15.0 0.0 31.6



#35

Dibromofluoromethane

Concen: 50.937 ug/l

RT: 8.165 min Scan# 1056

Delta R.T. -0.000 min

Lab File: VN084534.D

Acq: 28 Oct 2024 13:06




Tgt Ion:113 Resp: 95292


Ion Ratio Lower Upper

113 100

111 104.2 83.3 124.9

192 18.6 13.5 20.3









#50
Toluene-d8
Concen: 49.428 ug/l
RT: 10.565 min Scan# 1464
Delta R.T. -0.000 min
Lab File: VN084534.D
Acq: 28 Oct 2024 13:06




Instrument :
MSVOA_N
ClientSampleId :
VIBLK


Tgt Ion: 98 Resp: 340205
Ion Ratio Lower Upper
98 100
100 64.1 52.7 79.1



#62
4-Bromofluorobenzene
Concen: 46.573 ug/l
RT: 12.847 min Scan# 1852
Delta R.T. -0.000 min
Lab File: VN084534.D
Acq: 28 Oct 2024 13:06

Tgt Ion: 95 Resp: 116781
Ion Ratio Lower Upper
95 100
174 74.1 0.0 145.2
176 70.5 0.0 140.0









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.865 min Scan# 1
Delta R.T. -0.000 min
Lab File: VN084534.D
Acq: 28 Oct 2024 13:06

Instrument :
MSVOA_N
ClientSampleId :
VIBLK

Tgt Ion:117 Resp: 244760
Ion Ratio Lower Upper
117 100
82 56.2 47.2 70.8
119 32.5 25.4 38.0



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.788 min Scan# 2012
Delta R.T. -0.000 min
Lab File: VN084534.D
Acq: 28 Oct 2024 13:06

Tgt Ion:152 Resp: 100197
Ion Ratio Lower Upper
152 100
115 62.6 31.3 93.9
150 161.9 0.0 349.8

