

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102924\
 Data File : VN084554.D
 Acq On : 29 Oct 2024 10:59
 Operator : JC\MD
 Sample : VN1029WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1029WBL01

Quant Time: Oct 30 04:58:23 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 01 07:11:01 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.218	168	172818	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	299281	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	246662	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	92882	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	126644	49.425	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	98.840%
35) Dibromofluoromethane	8.165	113	101624	51.506	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	103.020%
50) Toluene-d8	10.565	98	346812	47.776	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	95.560%
62) 4-Bromofluorobenzene	12.847	95	112106	42.391	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	84.780%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102924\
Data File : VN084554.D
Acq On : 29 Oct 2024 10:59
Operator : JC\MD
Sample : VN1029WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1029WBL01

Quant Time: Oct 30 04:58:23 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 01 07:11:01 2024
Response via : Initial Calibration

