

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102924\
 Data File : VN084559.D
 Acq On : 29 Oct 2024 13:17
 Operator : JC\MD
 Sample : P4561-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BP-F-19

Quant Time: Oct 30 05:01:42 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 01 07:11:01 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.218	168	166163	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	294300	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	262850	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	112831	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	120394	48.868	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	97.740%
35) Dibromofluoromethane	8.165	113	96413	49.692	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.380%
50) Toluene-d8	10.565	98	359587	50.375	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	100.740%
62) 4-Bromofluorobenzene	12.847	95	124675	47.942	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	95.880%
Target Compounds						
					Qvalue	
16) Acetone	4.436	43	33136	31.309	ug/l #	87
18) Methyl Acetate	5.024	43	4338	1.881	ug/l #	87
20) Methylene Chloride	5.259	84	2897	1.347	ug/l	87
43) Isopropyl Acetate	8.688	43	130910	23.116	ug/l #	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102924\
Data File : VN084559.D
Acq On : 29 Oct 2024 13:17
Operator : JC\MD
Sample : P4561-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-F-19

Quant Time: Oct 30 05:01:42 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 01 07:11:01 2024
Response via : Initial Calibration

