

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102924\
 Data File : VN084565.D
 Acq On : 29 Oct 2024 15:41
 Operator : JC\MD
 Sample : P4578-06
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WB-305-BOT

Quant Time: Oct 30 05:04:42 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 01 07:11:01 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	153660	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	282987	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	251597	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	102583	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	115366	50.637	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	101.280%
35) Dibromofluoromethane	8.165	113	91913	49.267	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.540%
50) Toluene-d8	10.565	98	333504	48.589	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	97.180%
62) 4-Bromofluorobenzene	12.847	95	121123	48.438	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	96.880%
Target Compounds						
						Qvalue
16) Acetone	4.430	43	32247	32.949	ug/l	99
18) Methyl Acetate	5.024	43	5853	2.744	ug/l	92
20) Methylene Chloride	5.277	84	3038	1.527	ug/l #	78
43) Isopropyl Acetate	8.688	43	161001	29.566	ug/l #	80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102924\
Data File : VN084565.D
Acq On : 29 Oct 2024 15:41
Operator : JC\MD
Sample : P4578-06
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WB-305-BOT

Quant Time: Oct 30 05:04:42 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 01 07:11:01 2024
Response via : Initial Calibration

