

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102924\
 Data File : VN084567.D
 Acq On : 29 Oct 2024 16:29
 Operator : JC\MD
 Sample : P4574-06
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 GRAVEL-2

Quant Time: Oct 30 05:05:41 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 01 07:11:01 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.218	168	164873	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	305105	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	282635	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	124776	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	123328	50.450	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	100.900%
35) Dibromofluoromethane	8.165	113	100189	49.809	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.620%
50) Toluene-d8	10.565	98	365223	49.352	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.700%
62) 4-Bromofluorobenzene	12.847	95	138728	51.457	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	102.920%
Target Compounds						
						Qvalue
16) Acetone	4.436	43	34695	33.039	ug/l	94
18) Methyl Acetate	5.036	43	6894	3.012	ug/l	93
20) Methylene Chloride	5.271	84	2960	1.387	ug/l	87
43) Isopropyl Acetate	8.688	43	187936	32.011	ug/l #	77

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102924\
Data File : VN084567.D
Acq On : 29 Oct 2024 16:29
Operator : JC\MD
Sample : P4574-06
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
GRAVEL-2

Quant Time: Oct 30 05:05:41 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 01 07:11:01 2024
Response via : Initial Calibration

