

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN103024\
 Data File : VN084572.D
 Acq On : 30 Oct 2024 12:33
 Operator : JC\MD
 Sample : VSTDICC020
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICC020

Quant Time: Oct 31 18:19:07 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 31 18:18:08 2024
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By : Semsettin Yesilyurt 11/04/2024
 Supervised By : Mahesh Dadoda 11/04/2024

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.218	168	187576	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	319890	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	275857	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	130374	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.571	65	53102	19.600	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	39.200%#		
35) Dibromofluoromethane	8.165	113	41686	19.252	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	38.500%#		
50) Toluene-d8	10.565	98	155563	19.504	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	39.000%#		
62) 4-Bromofluorobenzene	12.847	95	57622	19.330	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	38.660%#		
Target Compounds					Qvalue	
2) Dichlorodifluoromethane	2.124	85	41443	19.219	ug/l	99
3) Chloromethane	2.359	50	54414	19.993	ug/l	97
4) Vinyl Chloride	2.512	62	46710	20.140	ug/l	93
5) Bromomethane	2.901	94	23244	18.916	ug/l	100
6) Chloroethane	3.042	64	30992	20.772	ug/l	94
7) Trichlorofluoromethane	3.442	101	76339	19.982	ug/l	96
8) Diethyl Ether	3.948	74	26904	19.963	ug/l	96
9) 1,1,2-Trichlorotrifluo...	4.342	101	42830	19.840	ug/l	99
10) Methyl Iodide	4.565	142	57315	19.865	ug/l	96
11) Tert butyl alcohol	5.559	59	33927	94.871	ug/l	99
12) 1,1-Dichloroethene	4.318	96	41980	19.653	ug/l	89
13) Acrolein	4.171	56	38007	106.582	ug/l	96
14) Allyl chloride	5.006	41	70420	20.328	ug/l	92
15) Acrylonitrile	5.718	53	105912	102.304	ug/l	97
16) Acetone	4.430	43	79971	100.481	ug/l	99
17) Carbon Disulfide	4.689	76	127521	19.385	ug/l	98
18) Methyl Acetate	5.024	43	71604	22.882	ug/l	98
19) Methyl tert-butyl Ether	5.789	73	135170	20.645	ug/l	100
20) Methylene Chloride	5.271	84	47467	19.921	ug/l	88
21) trans-1,2-Dichloroethene	5.771	96	44697	20.378	ug/l	93
22) Diisopropyl ether	6.671	45	144115	20.937	ug/l	97
23) Vinyl Acetate	6.594	43	490034	103.232	ug/l	100
24) 1,1-Dichloroethane	6.559	63	83555	20.218	ug/l	97
25) 2-Butanone	7.483	43	130541	103.281	ug/l	99
26) 2,2-Dichloropropane	7.483	77	73066	20.313	ug/l	99
27) cis-1,2-Dichloroethene	7.483	96	52279	20.412	ug/l	97
28) Bromochloromethane	7.806	49	29146	17.702	ug/l	91
29) Tetrahydrofuran	7.836	42	86837	103.560	ug/l	97
30) Chloroform	7.959	83	85675	20.368	ug/l	100
31) Cyclohexane	8.247	56	71698	19.169	ug/l	96
32) 1,1,1-Trichloroethane	8.165	97	78458	20.493	ug/l	97
36) 1,1-Dichloropropene	8.371	75	60208	20.050	ug/l	99
37) Ethyl Acetate	7.559	43	54942	20.165	ug/l	97
38) Carbon Tetrachloride	8.359	117	68076	20.282	ug/l	98
39) Methylcyclohexane	9.600	83	63305	20.722	ug/l	93
40) Benzene	8.600	78	193129	20.026	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN103024\
 Data File : VN084572.D
 Acq On : 30 Oct 2024 12:33
 Operator : JC\MD
 Sample : VSTDICC020
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICC020

Quant Time: Oct 31 18:19:07 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 31 18:18:08 2024
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By : Semsettin Yesilyurt 11/04/2024
 Supervised By : Mahesh Dadoda 11/04/2024

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.771	41	30596	20.882	ug/l	97
42) 1,2-Dichloroethane	8.665	62	63026	20.180	ug/l	99
43) Isopropyl Acetate	8.688	43	104951	20.277	ug/l	98
44) Trichloroethene	9.347	130	44176	19.862	ug/l	99
45) 1,2-Dichloropropane	9.618	63	45764	20.229	ug/l	93
46) Dibromomethane	9.706	93	32857	20.583	ug/l	99
47) Bromodichloromethane	9.888	83	67288	19.997	ug/l #	98
48) Methyl methacrylate	9.682	41	43398	20.466	ug/l	97
49) 1,4-Dioxane	9.700	88	15417	410.291	ug/l #	79
51) 4-Methyl-2-Pentanone	10.447	43	266010	102.417	ug/l	100
52) Toluene	10.629	92	117530	20.838	ug/l	99
53) t-1,3-Dichloropropene	10.835	75	68847	19.766	ug/l	99
54) cis-1,3-Dichloropropene	10.312	75	74498	20.208	ug/l	95
55) 1,1,2-Trichloroethane	11.012	97	42791	20.141	ug/l	95
56) Ethyl methacrylate	10.871	69	63347	20.613	ug/l	96
57) 1,3-Dichloropropane	11.159	76	74082	20.317	ug/l	100
58) 2-Chloroethyl Vinyl ether	10.159	63	146372	101.786	ug/l	98
59) 2-Hexanone	11.194	43	192335	101.735	ug/l	99
60) Dibromochloromethane	11.359	129	50039	20.665	ug/l	99
61) 1,2-Dibromoethane	11.465	107	43666	20.070	ug/l	100
64) Tetrachloroethene	11.106	164	36798	20.054	ug/l	92
65) Chlorobenzene	11.888	112	126753	20.539	ug/l	96
66) 1,1,1,2-Tetrachloroethane	11.959	131	43201	20.125	ug/l	99
67) Ethyl Benzene	11.965	91	212737	20.651	ug/l	99
68) m/p-Xylenes	12.071	106	162754	41.732	ug/l	99
69) o-Xylene	12.400	106	77396	21.214	ug/l	99
70) Styrene	12.412	104	133059	20.907	ug/l	96
71) Bromoform	12.576	173	32901	20.369	ug/l #	99
73) Isopropylbenzene	12.694	105	193023	21.127	ug/l	100
74) N-amyl acetate	12.494	43	80939	20.815	ug/l	98
75) 1,1,2,2-Tetrachloroethane	12.935	83	60645	19.886	ug/l	98
76) 1,2,3-Trichloropropane	12.994	75	49947m	19.414	ug/l	
77) Bromobenzene	12.976	156	48017	19.550	ug/l	97
78) n-propylbenzene	13.035	91	225085	21.024	ug/l	99
79) 2-Chlorotoluene	13.123	91	148533	21.233	ug/l	99
80) 1,3,5-Trimethylbenzene	13.171	105	161619	21.343	ug/l	100
81) trans-1,4-Dichloro-2-b...	12.735	75	22047	20.070	ug/l #	85
82) 4-Chlorotoluene	13.218	91	147989	20.102	ug/l	100
83) tert-Butylbenzene	13.435	119	138079	22.033	ug/l	99
84) 1,2,4-Trimethylbenzene	13.482	105	170246	21.783	ug/l	99
85) sec-Butylbenzene	13.618	105	188158	21.255	ug/l	100
86) p-Isopropyltoluene	13.729	119	156598	21.569	ug/l	99
87) 1,3-Dichlorobenzene	13.729	146	92310	19.258	ug/l	99
88) 1,4-Dichlorobenzene	13.812	146	92940	20.458	ug/l	99
89) n-Butylbenzene	14.053	91	136771	20.139	ug/l	99
90) Hexachloroethane	14.329	117	32199	19.887	ug/l	100
91) 1,2-Dichlorobenzene	14.106	146	91138	19.787	ug/l	98
92) 1,2-Dibromo-3-Chloropr...	14.717	75	11962	19.362	ug/l	96
93) 1,2,4-Trichlorobenzene	15.394	180	46698	18.523	ug/l	96
94) Hexachlorobutadiene	15.500	225	22216	20.066	ug/l	96
95) Naphthalene	15.635	128	153972	20.404	ug/l	99
96) 1,2,3-Trichlorobenzene	15.835	180	49699	20.308	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN103024\
Data File : VN084572.D
Acq On : 30 Oct 2024 12:33
Operator : JC\MD
Sample : VSTDICC020
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 11/04/2024
Supervised By :Mahesh Dadoda 11/04/2024

Quant Time: Oct 31 18:19:07 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 31 18:18:08 2024
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

