

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN103024\
 Data File : VN084577.D
 Acq On : 30 Oct 2024 15:06
 Operator : JC\MD
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :

Quant Time: Oct 31 18:28:48 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 31 18:24:48 2024
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By : Semsettin
 Yesilyurt

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) Pentafluorobenzene	8.224	168	192202	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	324264	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	291586	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	142817	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	132029	47.560	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	=	95.120%	
35) Dibromofluoromethane	8.165	113	104306	47.522	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	95.040%	
50) Toluene-d8	10.565	98	390989	48.359	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	=	96.720%	
62) 4-Bromofluorobenzene	12.847	95	144180	47.715	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	=	95.420%	
Target Compounds						
						Qvalue
2) Dichlorodifluoromethane	2.124	85	101605	45.985	ug/l	97
3) Chloromethane	2.359	50	118060	45.299	ug/l	99
4) Vinyl Chloride	2.512	62	110340	46.430	ug/l	95
5) Bromomethane	2.900	94	54764	43.494	ug/l	100
6) Chloroethane	3.071	64	68330	46.348	ug/l	93
7) Trichlorofluoromethane	3.471	101	178683	45.644	ug/l	98
8) Diethyl Ether	3.959	74	64407	46.640	ug/l	98
9) 1,1,2-Trichlorotrifluo...	4.359	101	101094	45.702	ug/l	98
10) Methyl Iodide	4.577	142	137594	46.542	ug/l	97
11) Tert butyl alcohol	5.536	59	80556	219.840	ug/l	99
12) 1,1-Dichloroethene	4.330	96	99235	45.338	ug/l	99
13) Acrolein	4.171	56	79788	218.362	ug/l	98
14) Allyl chloride	5.012	41	170293	47.975	ug/l	92
15) Acrylonitrile	5.718	53	251141	236.747	ug/l	100
16) Acetone	4.430	43	196070	244.445	ug/l	100
17) Carbon Disulfide	4.700	76	287697	42.682	ug/l	99
18) Methyl Acetate	5.024	43	132834	44.944	ug/l	99
19) Methyl tert-butyl Ether	5.789	73	326067	48.603	ug/l	97
20) Methylene Chloride	5.265	84	111618	45.716	ug/l	91
21) trans-1,2-Dichloroethene	5.777	96	101181	45.019	ug/l	98
22) Diisopropyl ether	6.671	45	343763	48.739	ug/l	99
23) Vinyl Acetate	6.600	43	1204384	247.614	ug/l	99
24) 1,1-Dichloroethane	6.559	63	196518	46.408	ug/l	97
25) 2-Butanone	7.483	43	307765	237.635	ug/l	97
26) 2,2-Dichloropropane	7.488	77	166234	45.103	ug/l	97
27) cis-1,2-Dichloroethene	7.483	96	121280	46.214	ug/l	97
28) Bromochloromethane	7.806	49	92741	54.971	ug/l	97
29) Tetrahydrofuran	7.835	42	211210	245.822	ug/l	98
30) Chloroform	7.965	83	203935	47.316	ug/l	100
31) Cyclohexane	8.253	56	172671	45.053	ug/l	99
32) 1,1,1-Trichloroethane	8.165	97	183488	46.774	ug/l	98
36) 1,1-Dichloropropene	8.365	75	141152	46.370	ug/l	100
37) Ethyl Acetate	7.559	43	132993	48.154	ug/l	99
38) Carbon Tetrachloride	8.359	117	157005	46.145	ug/l	97
39) Methylcyclohexane	9.600	83	154743	49.969	ug/l	99
40) Benzene	8.606	78	452351	46.273	ug/l	99

11/04/2024
 Supervised By : Mahesh
 Padoda

11/04/2024

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN103024\
 Data File : VN084577.D
 Acq On : 30 Oct 2024 15:06
 Operator : JC\MD
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :

Quant Time: Oct 31 18:28:48 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 31 18:24:48 2024
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By :Semsettin
 Yesilyurt

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.777	41	77176	51.964	ug/l	99
42) 1,2-Dichloroethane	8.671	62	149173	47.118	ug/l	99
43) Isopropyl Acetate	8.688	43	232558	45.492	ug/l	98
44) Trichloroethene	9.347	130	100849	44.732	ug/l	99
45) 1,2-Dichloropropane	9.618	63	108229	47.196	ug/l	96
46) Dibromomethane	9.706	93	74292	45.911	ug/l	96
47) Bromodichloromethane	9.888	83	158814	46.560	ug/l #	99
48) Methyl methacrylate	9.676	41	109745	51.057	ug/l	97
49) 1,4-Dioxane	9.694	88	40527	1063.992	ug/l #	95
51) 4-Methyl-2-Pentanone	10.447	43	665397	252.729	ug/l	98
52) Toluene	10.629	92	277957	48.616	ug/l	100
53) t-1,3-Dichloropropene	10.835	75	162100	45.911	ug/l	99
54) cis-1,3-Dichloropropene	10.312	75	175008	46.832	ug/l	99
55) 1,1,2-Trichloroethane	11.018	97	102553	47.618	ug/l	97
56) Ethyl methacrylate	10.871	69	164370	52.764	ug/l	98
57) 1,3-Dichloropropane	11.165	76	176546	47.765	ug/l	100
58) 2-Chloroethyl Vinyl ether	10.159	63	408174	280.013	ug/l	98
59) 2-Hexanone	11.194	43	489776	255.570	ug/l	99
60) Dibromochloromethane	11.359	129	120465	49.079	ug/l	99
61) 1,2-Dibromoethane	11.465	107	102316	46.393	ug/l	100
64) Tetrachloroethene	11.100	164	88124	45.436	ug/l	98
65) Chlorobenzene	11.888	112	292713	44.872	ug/l	95
66) 1,1,1,2-Tetrachloroethane	11.959	131	105108	46.323	ug/l	99
67) Ethyl Benzene	11.965	91	526140	48.319	ug/l	99
68) m/p-Xylenes	12.070	106	408029	98.979	ug/l	99
69) o-Xylene	12.400	106	194895	50.538	ug/l	99
70) Styrene	12.412	104	333668	49.600	ug/l	98
71) Bromoform	12.576	173	78822	46.166	ug/l #	98
73) Isopropylbenzene	12.694	105	491033	49.063	ug/l	100
74) N-ethyl acetate	12.494	43	209049	49.076	ug/l	99
75) 1,1,2,2-Tetrachloroethane	12.935	83	147181	44.058	ug/l	99
76) 1,2,3-Trichloropropane	12.994	75	122725m	42.988	ug/l	
77) Bromobenzene	12.976	156	119728	44.500	ug/l	97
78) n-propylbenzene	13.035	91	576495	49.156	ug/l	98
79) 2-Chlorotoluene	13.123	91	367294	47.930	ug/l	100
80) 1,3,5-Trimethylbenzene	13.170	105	420675	50.714	ug/l	100
81) trans-1,4-Dichloro-2-b...	12.735	75	54512	45.300	ug/l	99
82) 4-Chlorotoluene	13.223	91	361596	44.837	ug/l	98
83) tert-Butylbenzene	13.435	119	361800	52.701	ug/l	98
84) 1,2,4-Trimethylbenzene	13.482	105	428517	50.053	ug/l	100
85) sec-Butylbenzene	13.617	105	483212	49.829	ug/l	100
86) p-Isopropyltoluene	13.729	119	410129	51.568	ug/l	100
87) 1,3-Dichlorobenzene	13.729	146	219757	41.852	ug/l	99
88) 1,4-Dichlorobenzene	13.812	146	221608	45.904	ug/l	99
89) n-Butylbenzene	14.053	91	341775	45.941	ug/l	100
90) Hexachloroethane	14.329	117	77324	43.597	ug/l	98
91) 1,2-Dichlorobenzene	14.106	146	220824	43.765	ug/l	100
92) 1,2-Dibromo-3-Chloropr...	14.717	75	30895	45.651	ug/l	98
93) 1,2,4-Trichlorobenzene	15.388	180	111270	40.290	ug/l	98
94) Hexachlorobutadiene	15.500	225	47959	39.544	ug/l	97
95) Naphthalene	15.641	128	375360	45.408	ug/l	100
96) 1,2,3-Trichlorobenzene	15.841	180	108830	40.595	ug/l	99

11/04/2024
 Supervised By :Mahesh
 Dadoda

11/04/2024

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN103024\
Data File : VN084577.D
Acq On : 30 Oct 2024 15:06
Operator : JC\MD
Sample : VSTDICV050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :

Manual Integrations
APPROVED

Reviewed By :Semsettin
Yesilyurt

Quant Time: Oct 31 18:28:48 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 31 18:24:48 2024
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

(#) = qualifier out of range (m) = manual integration (+) = signals summed						

11/04/2024
Supervised By :Mahesh
Dadoda

11/04/2024

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN103024\
Data File : VN084577.D
Acq On : 30 Oct 2024 15:06
Operator : JC\MD
Sample : VSTDICV050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :

Quant Time: Oct 31 18:28:48 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 31 18:24:48 2024
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Semsettin
Yesilyurt

