

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN103024\
 Data File : VN084581.D
 Acq On : 30 Oct 2024 21:12
 Operator : JC\MD
 Sample : VN1030WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1030WBL01

Quant Time: Nov 01 04:54:57 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 31 18:45:38 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	181822	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	317923	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	282193	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	124291	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	130757	49.791	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	99.580%
35) Dibromofluoromethane	8.165	113	106498	49.489	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.980%
50) Toluene-d8	10.565	98	374287	47.216	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	94.440%
62) 4-Bromofluorobenzene	12.847	95	138531	46.760	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	93.520%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN103024\
Data File : VN084581.D
Acq On : 30 Oct 2024 21:12
Operator : JC\MD
Sample : VN1030WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1030WBL01

Quant Time: Nov 01 04:54:57 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 31 18:45:38 2024
Response via : Initial Calibration

