

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN103024\
 Data File : VN084591.D
 Acq On : 31 Oct 2024 01:21
 Operator : JC\MD
 Sample : P4597-09
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BLUE-2-1

Quant Time: Oct 31 18:32:12 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 31 18:24:48 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.218	168	182331	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	330565	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	301897	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	132253	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	133357	50.639	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	101.280%		
35) Dibromofluoromethane	8.165	113	107217	47.918	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	95.840%		
50) Toluene-d8	10.565	98	397359	48.210	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	96.420%		
62) 4-Bromofluorobenzene	12.847	95	148951	48.354	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	96.700%		
Target Compounds						
16) Acetone	4.430	43	19778	23.414	ug/l	96
20) Methylene Chloride	5.277	84	2409	1.040	ug/l #	66
43) Isopropyl Acetate	8.688	43	144121	27.269	ug/l #	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN103024\
Data File : VN084591.D
Acq On : 31 Oct 2024 01:21
Operator : JC\MD
Sample : P4597-09
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BLUE-2-1

Quant Time: Oct 31 18:32:12 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 31 18:24:48 2024
Response via : Initial Calibration

